

Regularized Emission Image Reconstruction Using Imperfect Side Information

Jeffrey A. Fessler*, Neal H. Clinthorne, and W. Leslie Rogers
Division of Nuclear Medicine, University of Michigan

ABSTRACT

The inadequacy of the maximum-likelihood criterion for emission image reconstruction has spurred the development of several regularization methods. Despite the spatial variance of medical images, most of the proposed methods are spatially invariant. This paper reports an investigation of a spatially-variant penalized-likelihood method for tomographic image reconstruction based on a weighted Gibbs penalty. The penalty weights are determined from structural side information, such as the locations of anatomical boundaries in high-resolution magnetic resonance images. Such side information will be imperfect in practice, and a simple simulation demonstrates the importance of accounting for the errors in boundary locations. We discuss methods for prescribing the penalty weights when the side information is noisy. Simulation results suggest that even imperfect side information is useful for guiding spatially-variant regularization.

I. INTRODUCTION

Many investigators have noted the inadequacy of the maximum-likelihood (ML) criterion for emission image reconstruction, and have proposed regularization techniques that stabilize the emission estimate. Most such methods are spatially invariant; however, the spatial variance typical of medical images argues for the use of spatially-variant reconstruction methods. This paper proposes using side information, such as the locations of anatomical boundaries obtained from magnetic resonance (MR) images, to control a spatially-variant penalized-likelihood method based on weighted Gibbs functions. An important feature of this method is that it can accommodate imperfect side information.

The method described is a synthesis of three recent advances in emission image reconstruction. The measurement model includes the effects of attenuation and accidental coincidences, the importance of which is shown in [1]. The use of spatially-variant weights for a Gibbs

penalty is analogous to the "weighted-splines" approach described in [2] for a Gaussian noise model. This regularizer leads to an optimality criterion with an analytically intractable M-step, so we apply a variant of the GEM [3] iterative method. These points are considered in detail in Section II.

The approach taken here is to use the side information to generate Gibbs penalty weights that are then held fixed for the duration of the estimation process. A more sophisticated approach might allow the weights to vary with iteration. One such method involves supplementing the emission intensity parameters with parameters representing a "line process" [4-8]. These powerful techniques have demonstrated the capability of accommodating large registration errors in computer simulations. However, since the number of unknown parameters is at least tripled over the case of fixed weights, they are necessarily more computationally expensive. Also, the convergence properties of the associated algorithms are less well understood than the weighted GEM algorithm we propose. We believe that as MR-PET registration methods improve, there will be a class of applications, particularly in neurological imaging, where the simpler approach described in this paper will be adequate.

The benefits of structural side information, such as might be derived from MR images, will clearly be task dependent. Therefore, in this paper we depart from the conventional global performance criteria, such as likelihood or global mean-square error, and focus on a specific local figure-of-merit: the accuracy of quantifying total uptake within a small region of interest (ROI) surrounded by regions of relatively higher activity. Since the results of such a study will be context dependent, we have chosen a context that has significance to clinical investigators at our institution: quantifying tracer uptake within the globus pallidus and the putamen for patients with Huntington's Disease from position emission tomographic (PET) measurements of regional benzodiazepine receptor density obtained by injection of [^{11}C] flumazenil [9, 10]. These small neurological structures (see Figure 8) are poorly quantified by conventional filtered back-projection images, due to spill over from the surrounding cortex. However, the boundaries of these structures are obtainable from MR

*This work was supported in part by NCI Training Grant 5 T32 CA 09015, a DOE Alexander Hollaender Postdoctoral Fellowship, and NIH Grant CA 54362.

images, so an iterative reconstruction method that can exploit such side information could be beneficial.

Section II describes the basic models and method, which we have applied to both a representative one-dimensional profile and to a realistic two-dimensional computer phantom with regional activities that correspond to autoradiographic data. Sections III and IV report the simulation results. Section V discusses the future directions of this research.

II. METHOD

A. Measurement model

Accurate quantification requires the use of an accurate measurement model. In particular, as shown in [1], one should account for the effects of attenuation and accidental coincidences in PET by including them in the measurement model, rather than by precorrecting the measurements. Precorrection destroys the Poisson nature of the measurements.

For simplicity, we adopt the voxelized object model, and denote the rate of activity in the b th voxel by λ_b , $b = 1, \dots, B$. The PET system consists of D detector pairs. Let p_{db} denote the point-spread function (PSF) of the PET system, i.e., p_{db} is the probability that, in the absence of attenuation, an event from the b th voxel is detected by the d th detector pair, conditioned on it being detected. Thus $\sum_{d=1}^D p_{db} = 1$. Let q_b denote the overall detection probability for an event originating in the b th voxel, in the absence of attenuation. Let μ_d denote the survival probability for the d th detector pair, i.e., the probability that both of an annihilation-produced pair of photons emitted along the d detector pair tube are detected (not attenuated). Let r_d denote the rate of accidental coincidences for the d th detector pair. Then if y_d denotes the number of events counted by the d th detector pair, we assume the y_d 's have independent Poisson distributions:

$$y_d \sim \text{Poisson}\left(T \cdot \left(\sum_{b=1}^B a_{db} \lambda_b + r_d\right)\right), \quad (1)$$

where T is the product of the scan time and correction factors such as that for radioactive decay, and $a_{db} = \mu_d p_{db} q_b$. For simplicity, we absorb T into λ_b and r_d .

B. Regularizing Penalty Function

Although one could use (1) to define an estimation method based on the ML criterion, the resulting estimates have excessive variance for the specific tasks we are considering, as we show in simulations below. By considering instead an optimality criterion that is the difference between the log-likelihood and a penalty function that discourages excessive variation between neighboring voxels, one can significantly reduce the variance with only a small increase in bias, thereby reducing the total RMS error. How much

bias is tolerable is clearly task dependent, and is a subject needing further investigation.

One's choice for the penalty function again will be task dependent. Our hypothesis was that the task of quantifying uptake within a small cold spot would benefit from a spatially-variant penalty function, so a weighted Gibbs function is a natural choice. Specifically, we consider the following penalized-likelihood estimate:

$$\hat{\lambda} = \arg \max_{\lambda} \Phi(\lambda)$$

$$\Phi(\lambda) = -\mathbf{1}' \mathbf{A} \lambda + \mathbf{y}' \log(\mathbf{A} \lambda + \mathbf{r}) - \alpha V(\lambda), \quad (2)$$

where $\mathbf{1}$ is the column vector of ones of length D , $\mathbf{A} = \{a_{db}\}$, $\lambda = [\lambda_1, \dots, \lambda_B]$, $\mathbf{r} = [r_1, \dots, r_D]$, and $\mathbf{y} = [y_1, \dots, y_D]$. The weighted Gibbs function V is defined by:

$$V(\lambda) = \frac{1}{2} \sum_{i,j} w_{i,j} (\lambda_i - \lambda_j)^2.$$

The weights $w_{i,j}$ control the influence of the penalty. If the presence of an anatomical boundary in an MR image implies that the activities in two neighboring voxels are likely to be disparate, then the corresponding weight should be set relatively small, so as to avoid penalizing the discrepancy. Such a scheme reduces the "edge artifact" of spatially-invariant regularization [11]. It is important to note that this weighting method does not force uniformity within anatomical regions.

In case of perfect side information, one would set $w_{i,j}$ to one for neighboring pixel pairs that do not straddle a boundary, and all other $w_{i,j}$'s to zero. In practice, side information will be imperfect due to noise in MR images, registration errors, and discrepancies between anatomical and functional boundaries. As we demonstrate empirically in Section III, it is essential to account for such errors. The weighted Gibbs penalty method lends itself naturally to the following approach: we first use the side information to generate weights that would be optimal if the boundaries were perfect, and then blur or dilate the weights with a kernel whose width corresponds to the uncertainty in the side information. Though this is not necessarily the optimal method for accommodating imperfect side information (cf [12]), it has yielded reasonable results in simulations.

C. Iterative Algorithm

Historically, the use of objective functions such as (2), with its Gibbs penalty, has been hampered by the slow convergence of the associated stochastic maximization procedures or by the uncertain behavior of locally convergent methods. Following the usual estimate-maximize (EM) algorithm derivation, one can show that the E step for (2) is:

$$\hat{y}_d(\lambda^i) = \sum_{b=1}^B a_{db} \lambda_b^i + r_d,$$

$$n_{db}^i = y_d a_{db} \lambda_b^i / \hat{y}_d(\lambda^i),$$

where λ^i denotes the emission estimate after the i th iteration, and n_{db}^i is the conditional expectation of the number of events in the d th detector due to the b th voxel. The M step requires maximizing:

$$\sum_{d=1}^D \sum_{b=1}^B [-a_{db} \lambda_b^{i+1} + n_{db}^i \log(a_{db} \lambda_b^{i+1})] - \alpha V(\lambda^{i+1}) \quad (3)$$

over λ^{i+1} . The resulting coupled set of equations appears to have no analytical solution. We experimented with the “one step late” (OSL) method of Green [13], but found the necessity of line-search operations [14] to be computationally prohibitive. We have adopted the generalized estimate-maximize (GEM) method of Hebert and Leahy [3,15], which, although originally applied to SPECT, is also applicable to the PET measurement model (1). Zeroing the derivative of (3) with respect to λ_b^{i+1} yields:

$$0 = - \sum_{d=1}^D a_{db} + \left(\sum_{d=1}^D n_{db}^i \right) / \lambda_b^{i+1} - \alpha \sum_j w_{b,j} (\lambda_b^{i+1} - \lambda_j^{i+1}). \quad (4)$$

The GEM approach is to first set $\lambda_b^{i+1} = \lambda_b^i$ for all b , then to loop through the b 's in some order and to replace λ_b^{i+1} with the unique positive root of (4). After considering the discussion in [16], we chose the following ordering for our one-dimensional simulations: on even iterations, the even voxels are updated first, and then the odd voxels are updated; the opposite order holds for odd iterations. Unlike the method in [16], this is guaranteed to increase the penalized likelihood each iteration [3]. For the two-dimensional simulations, we applied the raster-scan order of [3].

The convergence of the GEM algorithm has been addressed in [3,14], under the assumption that the penalized log-likelihood (2) is a strictly convex function of λ . This was established in [14] by showing that the log-likelihood (for a SPECT measurement model) is convex, and the penalty term is strictly convex. In our case, it is possible that several weights could be set to zero, in which case the penalty term may not be strictly convex (although it would remain convex). Fortunately, the presence of accidental coincidences in PET ensures that the log-likelihood term is strictly convex, provided the matrix \mathbf{A} has full column rank [17].

III. 1D SIMULATION

To explore the possible benefits of spatially-variant regularization in the presence of imperfect structural side information, we performed simulations based on the one-dimensional profile shown in Figure 1. This profile is representative of the flumazenil quantification task. The nonuniform cold spot represents the putamen, which is adjacent to the globus pallidus and the cortex, both of which have significantly higher activity. For this simulation, the diameter of the cold spot is 7 pixels. The system matrix \mathbf{A} corresponds to a triangular point spread function with

a FWHM of 5 pixels. A typical measurement realization \mathbf{y} is shown in Figure 2.

Our task is to quantify the total uptake within the cold spot. This task requires two components: 1) reconstructing the activity distribution from the noisy measurements, and 2) identifying the boundaries of the region of interest (ROI) and integrating the activity within that ROI. Here we focus only on the first task by integrating the activity within the *true* ROI (pixels 33 through 39) for all simulations.

We have examined the performance of the GEM reconstruction method as a function of the regularization parameter α in four scenarios. For each scenario and for each value of α , 50 measurement realizations were reconstructed via 1000 iterations of the GEM algorithm.

- Case 1: No side information was available, so all the weights w_b were set to 1. This corresponds to conventional spatially-invariant regularization.
- Case 2: Perfect side information corresponding to the edge locations was used. Specifically, $w_{32} = w_{39} = 0$, and all other weights were set to 1. Thus, the weights for pixel pairs that straddled an edge of the cold spot were set to 0, thereby reducing spill over.
- Case 3: Imperfect side information was simulated. Let b_l and b_r denote the left and right edges of the cold spot as they might be determined from an MR image. (These two values are the side information). For each realization, b_l was randomly selected from the set $\{31, 32, 33\}$, and b_r was randomly selected from the set $\{38, 39, 40\}$. For Case 3, the side information was applied “blindly,” i.e., we set $w_{b_l} = w_{b_r} = 0$ and all other weights to 1, despite the fact that b_l and b_r are usually incorrect.
- Case 4: Imperfect side information with the same error distribution as in Case 3. However, for Case 4 we made the following heuristic attempt to account for the errors in b_l and b_r : we set $w_{b_l-1} = w_{b_l} = w_{b_l+1} = w_{b_r-1} = w_{b_r} = w_{b_r+1} = 0.01$, and all other weights to 1. This small band of weights corresponds to a dilation of the ideal weights, and allows for a rapid activity transition in the neighborhood of edge location specified by the noisy side information.

Figure 3 displays the percent root mean-square (RMS) error in total uptake within the cold spot as a function of the regularization parameter α . The optimal performance of each method is summarized in Table I. It is useful to compare the mean of the 50 reconstructions for each case with the true activity distribution; these are shown in Figures 4–7. The dotted bands around each curve represent one standard deviation above and below the mean. The curve for each case corresponds to the value α_{opt} shown in Table I that minimized the RMS error. The four cases illustrate the tradeoff between bias and variance: Case 1 has high variance, but additional smoothing would cause more

increase in bias than decrease in variance. Case 2 has low variance and low bias since the edges are known perfectly. Case 4 improves over Case 3 by significantly reducing the bias with slight increase in variance.

IV. 2D SIMULATION

The one-dimensional simulation results encouraged our proceeding to evaluate the method on a realistic computer phantom. Figure 8 shows the true emission distribution used for this study. The anatomical boundaries in this image were obtained by manually tracing a photographic atlas. The regional emission intensities were assigned relative values based on autoradiographic measurements.

The simulated PET measurements included the effects of nonuniform attenuation (due to skull), accidental coincidences ($\approx 8\%$), and finite detector response (6mm FWHM). We assumed the survival probabilities and accidental coincidence rates were known. (However, our method has been implemented in conjunction with the joint-likelihood method of [18] that accounts for noise in transmission measurements). Each PET measurement realization contained approximately 10^6 total counts, distributed over 100 angles by 64 bins. The reconstructed object size was 50×64 pixels with 4mm sides.

For each reconstruction method shown in the figures, 30 pseudo-random Poisson distributed realizations were reconstructed. For each realization, the total uptake within the true region of support for the globus pallidus and putamen were computed. The conventional filtered back-projection reconstructions are shown in Figure 8, for both a ramp filter, and a Hanning-windowed ramp filter with a cutoff at the Nyquist frequency. The percent bias, standard deviation, and root mean-square (RMS) error are summarized in Table II. The iterative methods were ran for 300 iterations, with the total uptake computed for every fifth iteration. The resulting bias and variance are plotted as trajectories in Figures 10–12. The semi-ellipses in those figures correspond to contours of constant RMS error.

We compared three iterative methods: sieve-constrained EM algorithm [11], weighted Gibbs penalty GEM with perfect side information, and GEM with blurred weights. The latter case corresponds to degrading the perfect side information with a 6mm FWHM kernel—this figure is comparable to recently published MR-PET registration accuracies. Each method was evaluated for several values of the regularization parameter α . The sieve-constrained EM algorithm was evaluated for larger kernel sizes than 4mm FWHM, but the resulting uptake estimates had significantly worse RMS error. Sieve-constrained EM is a spatially invariant method, and therefore does not reduce spill over into the small ROI's considered here. Figure 9 displays representative reconstructions for the three iterative methods. Table II summarizes the error statistics for the different methods at the 300th iteration.

V. DISCUSSION

The results summarized in Tables I and II demonstrate that the use of structural side information, in conjunction with a spatially variant reconstruction method, can significantly reduce RMS error over spatially invariant regularization. In the one-dimensional case, the RMS error was reduced by almost a factor of 3 with the use of perfect side information. For the two-dimensional simulation, the RMS error in uptake within the globus pallidus was reduced by over a factor of 4. However, the one-dimensional results clearly demonstrate that if the side information is imperfect, then using it “blindly” is unlikely to be significantly beneficial for this type of quantification task. In the one-dimensional example, we applied a simple heuristic scheme (weight dilation) to account for the uncertainty in the side information, and were able to recover some of the benefits of the side information, despite its imperfections. The results with blurred weights for the two-dimensional case are less impressive, and more investigation into how to efficiently use imperfect side information is clearly needed.

VI. ACKNOWLEDGEMENT

The authors gratefully acknowledge the contributions of G. Hutchins and R. Koeppe.

REFERENCES

- [1] D. G. Politte and D. L. Snyder, “Corrections for accidental coincidences and attenuation in maximum-likelihood image reconstruction for positron-emission tomography,” *IEEE Transactions on Medical Imaging*, vol. 10, pp. 82–89, Mar. 1991.
- [2] X. Wang, B. G. Schunck, and W. L. Rogers, “A iterative reconstruction algorithm using spatially variant regularization for emission tomography,” *Journal of Nuclear Medicine (Abstract Book)*, vol. 32, pp. 936–937, May 1991.
- [3] T. Hebert and R. Leahy, “A generalized EM algorithm for 3-D Bayesian reconstruction from Poisson data using Gibbs priors,” *IEEE Transactions on Medical Imaging*, vol. 8, pp. 194–202, June 1989.
- [4] C. T. Chen, X. Ouyang, W. H. Wong, and X. Hu, “Improvement of PET image reconstruction using high-resolution anatomic images,” in *Conference Record of the 1991 IEEE Nuclear Science Symposium and Medical Imaging Conference*, p. 2062, 1991.
- [5] V. E. Johnson, W. H. Wong, X. Hu, and C. T. Chen, “Image restoration using Gibbs priors: Boundary modeling, treatment of blurring, and selection of hyperparameter,” *IEEE Transactions on Pattern Analysis and Machine Intelligence*, vol. 13, pp. 413–425, May 1991.
- [6] C. T. Chen, V. E. Johnson, W. H. Wong, X. Hu, and C. E. Metz, “Bayesian image reconstruction in positron emission tomography,” *IEEE Transactions on Nuclear Science*, vol. 37, pp. 636–641, Apr. 1990.

- [7] R. Leahy and X. H. Yan, "Statistical models and methods for PET image reconstruction," in *Proc. 1991 Amer. Stat. Assoc. Meeting*, 1991.
- [8] S. Geman and D. Geman, "Stochastic relaxation, Gibbs distributions, and Bayesian restoration of images," *IEEE Transactions on Pattern Analysis and Machine Intelligence*, vol. 6, pp. 721-741, Nov. 1984.
- [9] R. A. Koeppe, V. A. Holthoff, K. A. Frey, M. R. Kilbourn, and D. E. Kuhl, "Compartmental analysis of [^{11}C] flumazenil kinetics for the estimation of ligand transport rate and receptor distribution using positron emission tomography," *Journal of Cerebral Blood Flow and Metabolism*, vol. 11, no. 5, pp. 735-744, 1991.
- [10] V. A. Holthoff, R. A. Koeppe, K. A. Frey, J. B. Penney, D. S. Markel, D. E. Kuhl, and A. B. Young, "PET measures of benzodiazepine receptors in Huntington's disease," 1991. In review.
- [11] D. L. Snyder, M. I. Miller, L. J. Thomas, and D. G. Politte, "Noise and edge artifacts in maximum-likelihood reconstructions for emission tomography," *IEEE Transactions on Medical Imaging*, vol. 6, pp. 228-238, Sept. 1987.
- [12] P. C. Chiao, W. L. Rogers, A. O. Hero, and J. A. Fessler, "A lower bound with or without boundary side information for myocardial compartmental parameter estimation using PET or SPECT," *IEEE Transactions on Medical Imaging*, 1991. Submitted.
- [13] P. J. Green, "Bayesian reconstructions from emission tomography data using a modified EM algorithm," *IEEE Transactions on Medical Imaging*, vol. 9, pp. 84-93, Mar. 1990.
- [14] K. Lange, "Convergence of EM image reconstruction algorithms with Gibbs smoothing," *IEEE Transactions on Medical Imaging*, vol. 9, pp. 439-446, Dec. 1990. Corrections, June 1991 TMI.
- [15] T. Hebert and R. Leahy, "A Bayesian reconstruction algorithm for emission tomography using a Markov random field prior," in *Proc. SPIE 1092, Medical Imaging III: Image Processing*, pp. 458-4662, 1989.
- [16] G. T. Herman and D. Odhner, "Performance evaluation of an iterative image reconstruction algorithm for positron-emission tomography," *IEEE Transactions on Medical Imaging*, vol. 10, pp. 336-346, Sept. 1991.
- [17] J. A. Fessler, N. H. Clinthorne, and W. L. Rogers, "On complete-data spaces for PET reconstruction algorithms," 1991. Submitted to *IEEE Transactions on Medical Imaging*.
- [18] N. H. Clinthorne, J. A. Fessler, G. D. Hutchins, and W. L. Rogers, "Joint maximum likelihood estimation of emission and attenuation densities in PET," in *Conference Record of the 1991 IEEE Nuclear Science Symposium and Medical Imaging Conference*, vol. 3, pp. 1927-1932, 1991.

	α_{opt}	% RMS error
Case 1	10^{-4}	30.8 ± 1.9
Case 2	10^{-2}	11.0 ± 1.1
Case 3	10^{-3}	28.8 ± 2.1
Case 4	$10^{-2.5}$	17.4 ± 1.7

Table I: Percent RMS error for the optimal α for the four one-dimensional cases (see text).




Figure 1: Simulated one-dimensional activity distribution.




Figure 2: Measurement realization with 10^4 counts.




Figure 3: Percent RMS error for the four cases versus α . Case 1 (.), Case 2 (-), Case 3 (.), Case 4 (-).




Figure 4: True activity (-), mean reconstruction (o), and pointwise standard deviation (.) for Case 1.




Figure 6: True activity (-), mean reconstruction (o), and pointwise standard deviation (.) for Case 3.




Figure 5: True activity (-), mean reconstruction (o), and pointwise standard deviation (.) for Case 2.




Figure 7: True activity (-), mean reconstruction (o), and pointwise standard deviation (.) for Case 4.




Figure 8: Left: Simulated activity distribution. Middle: Filtered back-projection reconstruction with a ramp filter. Right: Filtered back-projection reconstruction with a Hanning window (cutoff at Nyquist frequency).




Figure 9: Left: Sieve-constrained EM reconstruction, FWHM=4mm. Middle: GEM reconstruction with perfect weights. $\alpha = e^{-2}$. Right: GEM reconstruction with blurred weights. $\alpha = e^{-7}$.

Method	Globus Pallidus			Putamen		
	bias	σ	RMS	bias	σ	RMS
FBP—Ramp	-47.0	6.2	47.4	-8.2	6.9	10.7
FBP—Hanning window	-56.2	3.6	56.3	-2.5	4.9	5.5
Sieve-constrained EM, 4mm	-39.6	9.1	40.6	-21.0	6.1	21.8
GEM, ideal weights, $\alpha = e^{-5}$	-14.4	14.8	20.7	-14.8	10.7	18.3
GEM, ideal weights, $\alpha = e^{-2}$	-3.3	11.2	11.7	-5.7	11.3	12.7
GEM, ideal weights, $\alpha = e^0$	-4.6	9.5	10.6	-8.5	10.5	13.5
GEM, blurred weights, $\alpha = e^{-5}$	-48.4	4.2	48.6	-24.7	4.6	25.1
GEM, blurred weights, $\alpha = e^{-7}$	-35.6	6.9	36.3	-18.0	9.1	20.2

Table II: Percent bias, standard deviation, and root mean-square (RMS) error for the total uptake within the two regions. The standard errors for these statistics are about $\pm 2\%$. The iterative methods show significant improvement for the smaller region (globus pallidus), but are unimpressive for the larger region (putamen).




Figure 10: Bias-vs-variance for 4mm sieve-constrained EM reconstructions. Left: globus pallidus. Right: putamen.




Figure 11: Bias-vs-variance for GEM reconstructions with ideal weights ($\alpha = e^{-2}$). Left: globus pallidus. Right: putamen.




Figure 12: Bias-vs-variance for GEM reconstructions with blurred weights ($\alpha = e^{-7}$). Left: globus pallidus. Right: putamen.