

Multi-Material Decomposition Using Statistical Image Reconstruction for Spectral CT

Yong Long* and Jeffrey A. Fessler, *Fellow, IEEE*

Abstract—Spectral computed tomography (CT) provides information on material characterization and quantification because of its ability to separate different basis materials. Dual-energy (DE) CT provides two sets of measurements at two different source energies. In principle, two materials can be accurately decomposed from DECT measurements. However, many clinical and industrial applications require three or more material images. For triple-material decomposition, a third constraint, such as volume conservation, mass conservation or both, is required to solve three sets of unknowns from two sets of measurements. The recently proposed flexible image-domain (ID) multi-material decomposition method assumes each pixel contains at most three materials out of several possible materials and decomposes a mixture pixel by pixel. We propose a penalized-likelihood (PL) method with edge-preserving regularizers for each material to reconstruct multi-material images using a similar constraint from sinogram data. We develop an optimization transfer method with a series of pixel-wise separable quadratic surrogate (PWSQS) functions to monotonically decrease the complicated PL cost function. The PWSQS algorithm separates pixels to allow simultaneous update of all pixels, but keeps the basis materials coupled to allow faster convergence rate than our previous proposed material- and pixel-wise SQS algorithms. Comparing with the ID method using 2-D fan-beam simulations, the PL method greatly reduced noise, streak and cross-talk artifacts in the reconstructed basis component images, and achieved much smaller root mean square errors.

Index Terms—Dual-energy computed tomography (CT), multi-material decomposition, optimization transfer, spectral CT, statistical image reconstruction.

I. INTRODUCTION

X-RAY computed tomography (CT) images the spatial distribution of attenuation coefficients of the object being scanned. Attenuation maps have many applications both in medical diagnosis and treatment and in industry for nondestructive evaluation. A conventional CT scanner measures a single sinogram at single X-ray source potential. Conventional image re-

construction methods process such measurements to produce a scalar-valued image of the scanned object.

In practice the scanned object always contains multiple materials. For example, the organs and tissues of human body under CT scans contain typical basis materials of blood, fat, muscle, water, cortical bone, air, and contrast agent [1], [2]. Material attenuation coefficients depend on the energy of the incident photons. An X-ray beam in clinical practice is usually composed of individual photons with a wide range of energies, and each photon energy is attenuated differently by the materials in the object. If uncorrected, this energy dependence causes artifacts in images reconstructed by conventional methods, such as beam-hardening artifacts [3]. This energy dependence also allows the possibility of basis-material decomposition [1], [4]–[8]. Numerous applications of two-material decomposition have been explored, including CT-based attenuation correction for positron emission tomography (PET) [7], [9], beam-hardening artifacts correction [10], [11], and virtual un-enhancement (VUE) CT [1], [8].

Dual-energy (DE) CT methods, pioneered by Alvarez and Macovski *et al.* [4], [12]–[15], are the most predominant approaches for reconstructing two basis materials (e.g., soft-tissue and bone). They decomposed the energy dependence of attenuation coefficients into two components, one approximated the photoelectric interaction and another approximated Compton scattering, and separated these two components from two sets of measurements at two different source energies. Although DECT methods were originally proposed in the late 1970s and early 1980s, DECT scanners became clinically available only recently with technological developments, such as fast kVp-switching, dual-source CT and dual-layer detectors. These new techniques have brought renewed interest in DECT [6], [7], [9], [16]–[26].

Several methods have been developed for reconstructing two basis materials from one CT scan with a single tube voltage setting. Ritchings and Pullan [27] described a technique for acquiring spectrally different data by filtering alternate detector elements. Taschereau *et al.* [28] retrofitted a preclinical microCT scanner with a filter wheel that alternates two beam filters between successive projections. One filter provides a low energy beam while the other filter provides a high energy beam. We [29] proposed a statistical penalized weighted least-squares (PWLS) method for reconstructing two basis materials from a single-voltage CT scan, exploiting the incident spectra difference of rays created by filtration, such as split [30] and bow-tie filters. One major limitation of decomposition methods based on single-voltage CT is the significant overlap in the two spectra that are generated by different filters.

Manuscript received February 09, 2014; revised April 02, 2014; accepted April 07, 2014. Date of publication April 25, 2014; date of current version July 30, 2014. This work is supported in part by the National Institutes of Health (NIH) under Grant R01 HL-098686 and Grant R01 CA-115870. *Asterisk indicates corresponding author.*

*Y. Long was with the CT Systems and Application Laboratory, GE Global Research Center, Niskayuna, NY 12309 USA. She is now with the University of Michigan–Shanghai Jiao Tong University Joint Institute, Shanghai Jiao Tong University, Shanghai 200240, China (e-mail: yong.long@sjtu.edu.cn).

J. A. Fessler is with the Department of Electrical Engineering and Computer Science, University of Michigan, Ann Arbor, MI 48109 USA (e-mail: fessler@umich.edu).

Color versions of one or more of the figures in this paper are available online at <http://ieeexplore.ieee.org>.

Digital Object Identifier 10.1109/TMI.2014.2320284

Many clinical and industrial applications desire three or more component images [1], [22], [31], [32]. Quantifying liver fat concentration requires images of four constitute materials, liver tissue, blood, fat, and contrast agent [1], [32]. Multi-material decomposition (MMD) can generate VUE images by removing the effect of contrast agents from contrast-enhanced CT exams without needing an additional contrast-free scan, reducing patient dose [1]. For radiotherapy, it is also useful to know the distributions of materials besides bone and soft tissue, such as calcium, metal (e.g., gold) and iodine-based contrast agent.

Typically, spectral CT methods reconstruct images of $L_0 = M_0$ basis materials from M_0 sets of measurements with M_0 different spectra [5]. Sukovic and Clinthorne [5] separated $L_0 = 3$ basis materials from $M_0 = 3$ sinograms acquired using $M_0 = 3$ distinct source voltages, one of which was near the K-edge of one basis material. Generally, spectral CT requires multiple scans or specialized scanner designs, such as quasi-monochromatic sources [33] or energy-resolved photon-counting detectors [34]. In this paper, we focus on MMD using DECT which is currently the only commercially available version of spectral CT. However, the general formulation is applicable to a variety of spectral CT approaches.

A third criteria, such as volume conservation [21], mass conservation [22] or both [1], can enable reconstructing three basis materials from DECT measurements. Volume (mass) conservation assumes the sum of the volumes (masses) of the three constituent materials is equivalent to the volume (mass) of the mixture. To reconstruct $L_0 > 3$ materials from DECT, solving this ill-posed problem requires additional assumptions. Mendonca *et al.* [1] proposed an image-domain (ID) method to reconstruct multiple materials pixel by pixel from a DECT scan. In addition to both volume and mass conservation assumptions, that method assumes that each pixel contains a mixture of at most three materials where the material types can vary between pixels. It establishes a material library containing all the possible triplets of basis materials for a specific application. It obtains a dual-material-density pair through projection-based decomposition approach from DECT measurements, and then generates a linear-attenuation-coefficient (LAC) pair for each pixel at two selected distinct energies (e.g., 70 and 140 keV). Given a LAC pair, a material triplet and the sum-to-one constraint that was derived using both the volume and mass conservation assumption, triple material decomposition is solved for each pixel. This method sequentially decomposes each pixel into different triplets in the material library in a prioritized order, and collects solutions of volume fractions that satisfy a box ($[0, 1]$) constraint and sum-to-one constraint. If there is a solution, it moves on to the next pixel and skips material triplets with lower priorities. If there is no feasible solution for all the material triplets, it relaxes the box constraints and accepts the volume fractions corresponding to the triplet with minimal Hausdorff distance to the LAC pair over all possible triplets.

Inspired by the ID method [1], [35] we proposed a penalized-likelihood (PL) method [36] with an edge-preserving regularizer to reconstruct multi-material images. It is well known that statistical image reconstruction methods based on physical models of the CT system and scanned object and statistical models of the measurements can obtain lower noise images with

higher quality. The proposed PL method considers each material image as a whole, instead of pixel-by-pixel, so prior knowledge, such as piecewise smoothness, can help solve the reconstruction problem.

The cost function of the PL method is minimized under the constraints that each pixel contains at most three materials, the volume fractions of basis materials sum to one, and the fractions are in the box $[0, 1]$. It is difficult to minimize the PL cost function directly. The preliminary PL method [36] applied optimization transfer principles to develop a series of pixel- and material-wise separable quadratic surrogates to monotonically decrease the cost function. The separability both in the pixel and material made the curvatures of surrogate functions large, causing slow convergence. In this paper, we propose an optimization transfer method with pixel-wise separable quadratic surrogates (PWSQS) that keep the materials coupled. The coupling in materials results in faster convergence. The maintained separability in pixels makes the PWSQS algorithm simultaneous and constrained optimization on each pixel easy.

We evaluated the proposed PL method on a modified NCAT chest phantom [37] containing fat, blood, omnipaque300 (iodine-based contrast agent), cortical bone, and air. Comparing with the ID method, the PL method was able to reconstruct component images with lower noise, greatly reduced streak artifacts, and alleviated the cross-talk phenomenon where a component of one material appearing in the image of another material. The rms errors of the PL method were about 60% lower for fat, blood, omnipaque300 and cortical bone, compared to the filtered ID method.

The organization of this paper is as follows. Section II introduces the physical models, including the measurement and object model. Section III describes the PL method. Section IV derives the PWSQS algorithm. Section V shows the results. Section VI presents conclusions.

II. PHYSICAL MODELS

A. Measurement Model

We use the following general model to describe the measurement physics for X-ray CT. The detector measures X-ray photons emerging from the object at $M_0 \geq 1$ different incident spectra. Based on current technologies, different incident spectra can be realized by either scanning with different X-ray spectra, such as fast kVp-switching [16] or dual-source CT [18], or by energy-resolved photon-counting detectors [34]. Let Y_{im} denote the measurement for the ray \mathcal{L}_{im} which is the i th ray for the m th energy scan, where $m = 1, \dots, M_0$, $i = 1, \dots, N_d$, and N_d is the number of rays. For notational simplicity we assume that the same number of rays are measured for each incident spectrum, but the physics model and methods presented in this paper can be easily generalized to cases where different incident spectra have different number of recorded rays. For a ray \mathcal{L}_{im} of infinitesimal width, the mean of the projection measurements can be expressed as

$$\bar{y}_{im} \triangleq \int I_{im}(\mathcal{E}) \exp\left(-\int_{\mathcal{L}_{im}} \mu(\vec{x}, \mathcal{E}) d\ell\right) d\mathcal{E} + r_{im} \quad (1)$$

where $\mu(\vec{x}, \mathcal{E})$ denotes the 3-D unknown spatially- and energy-dependent attenuation distribution, $\int_{\mathcal{L}_{im}} \cdot d\ell$ denotes the ‘‘line integral’’ function along line \mathcal{L}_{im} , and the incident X-ray intensity $I_{im}(\mathcal{E})$ incorporates the source spectrum and the detector gain. In reality, the measurements suffer from background signals such as Compton scatter, dark current, and noise. The ensemble mean of those effects (for the ray \mathcal{L}_{im}) is denoted as r_{im} . We treat each $I_{im}(\mathcal{E})$ and r_{im} as known nonnegative quantities. In practice, $I_{im}(\mathcal{E})$ can be determined by careful calibration [38], and r_{im} are estimated by some preprocessing steps prior to iterative reconstruction [39]–[41].

B. Object Model for Basis Material Decomposition

We describe the object model for basis material decomposition as

$$\mu(\vec{x}, \mathcal{E}) = \sum_{l=1}^{L_0} \mu_l(\mathcal{E}) \sum_{j=1}^{N_p} b_j(\vec{x}) x_{lj} \quad (2)$$

where $\mu_l(\mathcal{E})$ denotes the energy-dependent LAC of the l th material type, $b_j(\vec{x})$ denotes spatial basis functions (e.g., pixels), and x_{lj} denotes the volume fraction of the l th material in the j th pixel. Conventionally, one reconstructs $L_0 = M_0$ sets of basis materials from M_0 sets of measurements with M_0 different spectra [42].

Volume conservation [21], mass conservation [22] and combinations [1] have been used to provide extra information for solving $L_0 = M_0 + 1$ sets of unknowns from M_0 sets of independent measurements. Volume (mass) conservation assumes the sum of the volumes (masses) of the three constituent materials is equivalent to the volume (mass) of the mixture. Mendonca *et al.* [1] pointed out that any reasonable method for material decomposition already makes an implicit assumption of mass conservation. They used both volume and mass conservation to produce a model for the LAC of a mixture of materials. In this paper, we adopt their model where the volume fractions x_{lj} should satisfy the following sum to one and box constraints:

$$\begin{cases} \sum_{l=1}^{L_0} x_{lj} = 1, & \forall j \\ a_l \leq x_{lj} \leq b_l, & \forall l, j. \end{cases} \quad (3)$$

We relax the lower limit of the box constraint to be slightly smaller than 0, and the upper limit to be slightly greater than 1, i.e., $a_l = -\epsilon_l^a$ where $\epsilon_l^a \ll 1/2$, and $b_l = 1 + \epsilon_l^b$ where $\epsilon_l^b \ll 1/2$. This relaxation is similar to the work in [43], where negative values are allowed for the reconstructed densities of basis materials, such as water and iodine. This is because the estimated volume fractions are just coefficients for combining the linear attenuation coefficients of basis materials to produce the equivalent attenuation of a mixture. The sum to one constraint in (3) provides an extra criteria for solving $L_0 = M_0 + 1$ sets of unknowns from M_0 sets of measurements.

Additional assumptions are needed to estimate $L_0 > M_0 + 1$ sets of unknowns from M_0 sets of measurements [1]. We assume that each pixel contains at most $(M_0 + 1)$ types of materials and the material types can vary between pixels, i.e.,

$$\sum_{l=1}^{L_0} \mathbb{1}_{\{x_{lj} \neq 0\}} \leq M_0 + 1, \forall j. \quad (4)$$

Let Ω be a $(M_0 + 1)$ -tuple library containing all tuples formed from L_0 pre-selected materials of interest. Given a tuple ω in Ω , there are only $(M_0 + 1)$ unknowns for each pixel, which are solvable from M_0 sets of measurements with the help of the box and sum-to-one constraints given in (3). Note that air must be included as one basis material type even if it is typically not of primary interest. This is because there are always locations with LACs of zeros in the field-of-view (FOV) of the scanner and only the LAC of air is zero.

C. Combining Measurement and Object Model

Let \mathbf{x} denote the image vector $\mathbf{x} = (\mathbf{x}_1, \dots, \mathbf{x}_l, \dots, \mathbf{x}_{L_0}) \in \mathbb{R}^{N_p \times L_0}$ for $\mathbf{x}_l = (x_{l1}, \dots, x_{lj}, \dots, x_{lN_p}) \in \mathbb{R}^{N_p}$ of the l th material. Combining the general measurement model (1) and the object model (2), the mean of the projection measurement $\bar{y}_{im}(\mathbf{x})$ can be represented as follows:

$$\bar{y}_{im}(\mathbf{x}) = z_{im}(\mathbf{s}_{im}(\mathbf{x})) + r_{im} \quad (5)$$

for $m = 1, \dots, M_0$ and $i = 1, \dots, N_d$ where

$$z_{im}(\mathbf{s}_{im}(\mathbf{x})) \triangleq \int I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}(\mathbf{x})} d\mathcal{E}. \quad (6)$$

The linear attenuation vector $\boldsymbol{\mu}(\mathcal{E})$ and the sinogram vector $\mathbf{s}_{im}(\mathbf{x})$ are defined as

$$\boldsymbol{\mu}(\mathcal{E}) \triangleq (\mu_1(\mathcal{E}), \dots, \mu_{L_0}(\mathcal{E})) \quad (7)$$

$$\mathbf{s}_{im}(\mathbf{x}) \triangleq (s_{im1}(\mathbf{x}), \dots, s_{imL_0}(\mathbf{x})) \quad (8)$$

$$s_{iml}(\mathbf{x}) \triangleq [\mathbf{A}_m \mathbf{x}]_i = \sum_{j=1}^{N_p} a_{mij} x_{lj} \quad (9)$$

where \mathbf{A}_m denotes the $N_d \times N_p$ system matrix with entries

$$a_{mij} \triangleq \int_{\mathcal{L}_{im}} b_j(\vec{x}) d\ell. \quad (10)$$

As usual, we ignore the exponential edge gradient effects caused by the nonlinearity of Beer’s law [44], [45].

III. PENALIZED-LIKELIHOOD (PL) RECONSTRUCTION

For the case of normal clinical exposures, the X-ray CT measurements are often modeled as independently Poisson random variables with means (1), i.e.,

$$Y_{im} \sim \text{Poisson} \{ \bar{y}_{im}(\mathbf{x}) \}.$$

The corresponding negative log-likelihood for independent measurements Y_{im} has the form

$$\bar{L}(\mathbf{x}) \equiv \sum_{m=1}^{M_0} \sum_{i=1}^{N_d} t_{im}(\mathbf{s}_{im}(\mathbf{x})) \quad (11)$$

where \equiv means “equal to within irrelevant constants independent of \mathbf{x} ,” and

$$t_{im}(\mathbf{s}_{im}) \triangleq \bar{y}_{im}(\mathbf{s}_{im}) - Y_{im} h_{im}(\mathbf{s}_{im}) \quad (12)$$

$$h_{im}(\mathbf{s}_{im}) \triangleq \log(\bar{y}_{im}(\mathbf{s}_{im})). \quad (13)$$

We estimate component fraction images \mathbf{x} from the noisy measurements Y_{im} by minimizing a Penalized-Likelihood (PL) cost function subject to constraints given in (3) and (4) on the elements of \mathbf{x} as follows:

$$\hat{\mathbf{x}} = \arg \min_{\mathbf{x} \text{ subject to (3)\&(4)}} \Psi(\mathbf{x}) \quad (14)$$

$$\Psi(\mathbf{x}) \triangleq \bar{L}(\mathbf{x}) + R(\mathbf{x}). \quad (15)$$

The edge-preserving regularization term $R(\mathbf{x})$ is

$$R(\mathbf{x}) = \sum_{l=1}^{L_0} \beta_l R_l(\mathbf{x}_l) \quad (16)$$

where the regularizer for the l th material is

$$R_l(\mathbf{x}_l) = \sum_{j=1}^{N_p} \sum_{k \in \mathcal{N}_{l_j}} \kappa_{lj} \kappa_{lk} \psi_l(x_{lj} - x_{lk}) \quad (17)$$

where the potential function ψ_l is a hyperbola

$$\psi_l(t) = \frac{\delta_l^2}{3} \left(\sqrt{1 + 3 \left(\frac{t}{\delta_l} \right)^2} - 1 \right) \quad (18)$$

and where κ_{lj} and κ_{lk} are parameters encouraging uniform spatial resolution [46] and \mathcal{N}_{l_j} is some neighborhood of voxel x_{lj} . The regularization parameters β_l and δ_l can be chosen differently for different materials according to their properties.

Since the LAC of air is zero there is no contribution of air component to the data fitting term $\bar{L}(\mathbf{x})$ in (11), but the regularizer term $R(\mathbf{x})$ in (16) should include the air component because its image is piecewise smooth like other components. One could generalize the regularizer to consider joint sparsity of the component images [47], [48].

IV. OPTIMIZATION ALGORITHM

Because the cost function $\Psi(\mathbf{x})$ in (15) is difficult to minimize directly, we apply optimization transfer principles (OTP) [49]–[52] to develop an algorithm that monotonically decreases $\Psi(\mathbf{x})$ each iteration. We find a pixel-wise separable quadratic surrogate (PWSQS) $\phi^{(n)}(\mathbf{x})$ of the cost function, and then minimize $\phi^{(n)}(\mathbf{x})$ under constraints given in (3) and (4) on each

pixel. We loop over all tuples in the pre-determined material library Ω , minimize the surrogates under box and sum-to-one constraints in (3), and determine the optimal tuple for each pixel as the one minimizing the surrogate of that pixel.

A. Optimization Transfer Principles

The optimization transfer method [49], [50] replaces the cost function $\Psi(\mathbf{x})$ that is difficult to minimize with a surrogate function $\phi^{(n)}(\mathbf{x})$ that is easier to minimize at the n th iteration. The next estimate $\mathbf{x}^{(n+1)}$ is the minimizer of the surrogate function, i.e.,

$$\mathbf{x}^{(n+1)} \triangleq \arg \min_{\mathbf{x}} \phi^{(n)}(\mathbf{x}). \quad (19)$$

Repeatedly choosing a surrogate function and minimizing it at each iteration, one obtains a sequence of vectors $\{\mathbf{x}^{(n)}\}$ that monotonically decrease the original cost function $\Psi(\mathbf{x})$. The monotonicity is guaranteed by the following surrogate conditions:

$$\begin{aligned} \phi^{(n)}(\mathbf{x}^{(n)}) &= \Psi(\mathbf{x}^{(n)}) \\ \phi^{(n)}(\mathbf{x}) &\geq \Psi(\mathbf{x}), \quad \forall \mathbf{x} \in \mathbb{R}^{N_p \times L_0}. \end{aligned} \quad (20)$$

To derive surrogate functions for $\Psi(\mathbf{x})$ in (15), we consider the data fidelity term (11) and regularizer term (16) separately.

B. Surrogate of the Data Fidelity Term

1) *First Surrogate: Nonseparable Convex Surrogate:* The first step is to derive a convex surrogate as a function of the sino-gram vector \mathbf{s}_{im} . Since $h_{im}(\mathbf{s}_{im})$ is convex (see Appendix A), it is bounded below by its tangent plane

$$h_{im}(\mathbf{s}_{im}) \geq h_{im}(\mathbf{s}_{im}^{(n)}) + \nabla_{\mathbf{s}_{im}} h_{im}(\mathbf{s}_{im}^{(n)}) (\mathbf{s}_{im} - \mathbf{s}_{im}^{(n)}) \quad (21)$$

where $\mathbf{s}_{im}^{(n)} \triangleq \mathbf{s}_{im}(\mathbf{x}^{(n)})$ and $\nabla_{\mathbf{s}_{im}}$ denotes the row gradient with respect to \mathbf{s}_{im} . This inequality and the nonnegativity of Poisson random variable Y_{im} lead to our first surrogate for the data fidelity term $\bar{L}(\mathbf{x})$

$$\bar{L}(\mathbf{x}) \leq L_1^{(n)}(\mathbf{x}) \triangleq \sum_{m=1}^{M_0} \sum_{i=1}^{N_d} f_{1,im}^{(n)}(\mathbf{s}_{im}(\mathbf{x})) \quad (22)$$

where

$$\begin{aligned} f_{1,im}^{(n)}(\mathbf{s}_{im}) &\triangleq \bar{y}_{im}(\mathbf{s}_{im}) \\ &- Y_{im} \left(h_{im}(\mathbf{s}_{im}^{(n)}) + \nabla_{\mathbf{s}_{im}} h_{im}(\mathbf{s}_{im}^{(n)}) (\mathbf{s}_{im} - \mathbf{s}_{im}^{(n)}) \right). \end{aligned} \quad (23)$$

If the source is monoenergetic and if $r_{im} = 0$, then h_{im} is linear and the first surrogate $L_1^{(n)}(\mathbf{x})$ exactly equals the original negative log-likelihood data fidelity term $\bar{L}(\mathbf{x})$, so it should be a “reasonably tight” surrogate function.

2) *Second Surrogate: Nonseparable Quadratic Surrogate:* The second step is to find a quadratic surrogate of the first

convex surrogate given in (22) and (23). We rewrite $f_{1,im}^{(n)}(\mathbf{s}_{im})$ as follows:

$$\begin{aligned}
f_{1,im}^{(n)}(\mathbf{s}_{im}) &= \bar{y}_{im}(\mathbf{s}_{im}) - Y_{im} \nabla_{\mathbf{s}_{im}} h_{im}(\mathbf{s}_{im}^{(n)}) \mathbf{s}_{im} + C \\
&= \bar{y}_{im}(\mathbf{s}_{im}) - \frac{Y_{im}}{\bar{y}_{im}(\mathbf{s}_{im}^{(n)})} \nabla_{\mathbf{s}_{im}} \bar{y}_{im}(\mathbf{s}_{im}^{(n)}) \mathbf{s}_{im} + C \\
&= C + \int I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}} d\mathcal{E} \\
&\quad + \int Y_{im}^{(n)} I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}} \boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im} d\mathcal{E} \\
&= \int I_{im}(\mathcal{E}) f\left(\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}; Y_{im}^{(n)} e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}}\right) d\mathcal{E} + C
\end{aligned} \tag{24}$$

where $Y_{im}^{(n)} \triangleq Y_{im} / \bar{y}_{im}(\mathbf{s}_{im}^{(n)})$, and

$$f(x; \alpha) \triangleq e^{-x} + \alpha x \tag{25}$$

and C denotes the constants in (23) that are irrelevant to \mathbf{s}_{im} .

Let $\check{c}(\cdot; \alpha)$ given in (55) denote the optimal curvature for $f(x; \alpha)$ given in [52, Eq. (28)]. We form the optimal quadratic surrogate of $f(x; \alpha)$ using $\check{c}(\cdot; \alpha)$, and substitute it into (24) to obtain a quadratic surrogate of $f_{1,im}^{(n)}(\mathbf{s}_{im})$ as follows (see Appendix B):

$$\begin{aligned}
f_{2,im}^{(n)}(\mathbf{s}_{im}) &= f_{1,im}^{(n)}(\mathbf{s}_{im}^{(n)}) \\
&\quad + \nabla_{\mathbf{s}_{im}} f_{1,im}^{(n)}(\mathbf{s}_{im}^{(n)}) (\mathbf{s}_{im} - \mathbf{s}_{im}^{(n)}) \\
&\quad + \frac{1}{2} (\mathbf{s}_{im} - \mathbf{s}_{im}^{(n)})' \check{C}_{im}^{(n)} (\mathbf{s}_{im} - \mathbf{s}_{im}^{(n)})
\end{aligned} \tag{26}$$

where the $L_0 \times L_0$ curvature matrices are given by

$$\check{C}_{im}^{(n)} \triangleq \int I_{im}(\mathcal{E}) \check{c}_{im}^{(n)}(\mathcal{E}) \boldsymbol{\mu}(\mathcal{E}) \boldsymbol{\mu}'(\mathcal{E}) d\mathcal{E} \tag{27}$$

$$\check{c}_{im}^{(n)}(\mathcal{E}) \triangleq \check{c}\left(\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}; Y_{im}^{(n)} e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}}\right). \tag{28}$$

Summing this quadratic surrogate leads to the following nonseparable quadratic surrogate for the first convex surrogate $L_1^{(n)}(\mathbf{x})$:

$$L_1^{(n)}(\mathbf{x}) \leq L_2^{(n)}(\mathbf{x}) \triangleq \sum_{m=1}^{M_0} \sum_{i=1}^{N_d} f_{2,im}^{(n)}(\mathbf{s}_{im}(\mathbf{x})) \tag{29}$$

where $f_{2,im}^{(n)}(\mathbf{s}_{im}(\mathbf{x}))$ is given in (26).

3) *Third Surrogate: Pixel-Wise Separable Quadratic Surrogate:* Define pixel vector as $\mathbf{x}_j \triangleq (x_{1j}, \dots, x_{L_0j})$. The surrogate function $L_2^{(n)}(\mathbf{x})$ is a nonseparable quadratic function of \mathbf{x}_j . Nonseparable surrogates are inconvenient for simultaneous update algorithms and for enforcing the constraint in (3) and (4) on each pixel. To derive a simple simultaneous update algorithm that is fully parallelizable and suitable for ordered-subsets implementation [53], [54], we find next a pixel-wise separable

quadratic surrogate $L_3^{(n)}(\mathbf{x})$ of the nonseparable quadratic surrogate $L_2^{(n)}(\mathbf{x})$ by applying De Pierro's additive convexity trick [50], [53]. This novel pixel-wise separable quadratic surrogate remains nonseparable with respect to the basis materials.

Appendix C derives $L_3^{(n)}(\mathbf{x})$ in detail. It is defined as

$$L_2^{(n)}(\mathbf{x}) \leq L_3^{(n)}(\mathbf{x}) \triangleq \sum_{j=1}^{N_p} f_{3,j}^{(n)}(\mathbf{x}_j) \tag{30}$$

where

$$f_{3,j}^{(n)}(\mathbf{x}_j) \triangleq \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} \pi_{mij} f_{2,im}^{(n)}(b_{mij}(\mathbf{x}_j)) \tag{31}$$

$$b_{mij}(\mathbf{x}_j) \triangleq \frac{a_{mij}}{\pi_{mij}} (\mathbf{x}_j - \mathbf{x}_j^{(n)}) + \sum_{j=1}^{N_p} a_{mij} \mathbf{x}_j^{(n)}. \tag{32}$$

The π_{mij} values are non-negative and are zero only when a_{mij} is zero, and satisfy $\sum_{j=1}^{N_p} \pi_{mij} = 1$. For our empirical results, we use the following typical choice for π_{mij} [53]:

$$\pi_{mij} \triangleq \frac{|a_{mij}|}{a_{mi}}, \quad a_{mi} \triangleq \sum_{j=1}^{N_p} |a_{mij}|. \tag{33}$$

4) *Pixel-Wise Separable Quadratic Surrogate in Matrix-Vector Formation:* We have designed three surrogate functions sequentially having relationships

$$\bar{L}(\mathbf{x}) \leq L_1^{(n)}(\mathbf{x}) \leq L_2^{(n)}(\mathbf{x}) \leq L_3^{(n)}(\mathbf{x}). \tag{34}$$

Therefore, $L_3^{(n)}(\mathbf{x})$ is a surrogate function of $\bar{L}(\mathbf{x})$. It is much easier to minimize the surrogate $L_3^{(n)}$ under the proposed constraints since it is quadratic and separable with respect to pixels.

Combining the function value (60), gradient (61) and Hessian (62) derived in Appendix C, the pixel-wise separable quadratic surrogate has the following matrix-vector form:

$$\begin{aligned}
L_3^{(n)}(\mathbf{x}) &= \bar{L}(\mathbf{x}^{(n)}) + \nabla \bar{L}(\mathbf{x}^{(n)}) (\mathbf{x} - \mathbf{x}^{(n)}) \\
&\quad + \frac{1}{2} (\mathbf{x} - \mathbf{x}^{(n)})' \mathbf{D}_{\bar{L}}^{(n)} (\mathbf{x} - \mathbf{x}^{(n)})
\end{aligned} \tag{35}$$

where $\mathbf{D}_{\bar{L}}^{(n)}$ is a block diagonal matrix over $j = 1, \dots, N_p$, i.e.,

$$\mathbf{D}_{\bar{L}}^{(n)} \triangleq \text{diag} \left\{ \mathbf{D}_{\bar{L},j}^{(n)} \right\} \tag{36}$$

$$\mathbf{D}_{\bar{L},j}^{(n)} \triangleq \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} \frac{a_{mij}^2}{\pi_{mij}} \check{C}_{im}^{(n)} \tag{37}$$

and $\check{C}_{im}^{(n)}(\mathcal{E})$ is defined in (27). The $L_0 \times L_0$ matrix $\mathbf{D}_{\bar{L},j}^{(n)}$ is not diagonal due to the outer products in (27).

C. Surrogate of the Penalty Term

To derive a SQS function for the penalty term, we apply De Pierro's additive convexity trick [50], [53], [55] in a similar fashion and use Huber's optimal curvature [56, p. 185] for

the potential function ψ_l . The SQS function has the following matrix-vector form:

$$R^{(n)}(\mathbf{x}) = R(\mathbf{x}^{(n)}) + \nabla R(\mathbf{x}^{(n)}) (\mathbf{x} - \mathbf{x}^{(n)}) + \frac{1}{2} (\mathbf{x} - \mathbf{x}^{(n)})' \mathbf{D}_R^{(n)} (\mathbf{x} - \mathbf{x}^{(n)}) \quad (38)$$

where $\mathbf{D}_R^{(n)}$ is a diagonal matrix, i.e.,

$$\mathbf{D}_R^{(n)} \triangleq \text{diag} \left\{ \mathbf{D}_{R,j}^{(n)} \right\} \quad (39)$$

$$\mathbf{D}_{R,j}^{(n)} \triangleq \text{diag} \left\{ 4\beta_l \sum_{k \in \mathcal{N}_{l,j}} \kappa_{lj} \kappa_{lk} \omega_{\psi} \psi_l(x_{lj}^{(n)} - x_{lk}^{(n)}) \right\} \quad (40)$$

where $\omega_{\psi}(t) \triangleq \dot{\psi}(t)/t$.

D. Pixel-Wise Separable Quadratic Surrogate (PWSQS)

Combining the surrogates for the data fidelity term (35) and penalty term (38), the PWSQS function for the cost function $\Psi(\mathbf{x})$ is

$$\phi^{(n)}(\mathbf{x}) = L_3^{(n)}(\mathbf{x}) + R^{(n)}(\mathbf{x}). \quad (41)$$

It is easier to minimize $\phi^{(n)}(\mathbf{x})$ than the original cost function $\Psi(\mathbf{x})$ because it is quadratic and separable with respect to pixels by construction. For the sake of optimization under constraints on each pixel, we rewrite $\phi^{(n)}(\mathbf{x})$ in terms of j as

$$\phi^{(n)}(\mathbf{x}) \equiv \sum_{j=1}^{N_p} \phi_j^{(n)}(\mathbf{x}_j) \quad (42)$$

where

$$\phi_j^{(n)}(\mathbf{x}_j) \equiv \nabla_{\mathbf{x}_j} \Psi(\mathbf{x}^{(n)}) \mathbf{x}_j + \frac{1}{2} \left\| \mathbf{x}_j - \mathbf{x}_j^{(n)} \right\|_{\mathbf{D}_j^{(n)}}^2 \quad (43)$$

and the $L_0 \times L_0$ Hessian matrix of $\phi_j^{(n)}$ is

$$\mathbf{D}_j^{(n)} \triangleq \mathbf{D}_{L,j}^{(n)} + \mathbf{D}_{R,j}^{(n)}. \quad (44)$$

E. Optimization With Material Constraints

After designing the surrogate (42), the next step of the optimization transfer algorithm is to minimize $\phi^{(n)}(\mathbf{x})$ under constraints given in (3) and (4) on each pixel. Because $\phi^{(n)}(\mathbf{x})$ is pixel-wise separable, one can minimize $\phi_j^{(n)}(\mathbf{x}_j)$ for all pixels simultaneously. We now focus on the problem of minimizing $\phi_j^{(n)}(\mathbf{x}_j)$ under the proposed constraints.

Let ω denote a tuple in the material library Ω , i.e., $\omega \in \Omega$. The optimization problem on the j th pixel is

$$\begin{aligned} (\hat{\mathbf{x}}_j, \hat{\omega}) &= \arg \min_{\mathbf{x}_j, \omega \in \Omega} \phi_j^{(n)}(\mathbf{x}_j) \\ \text{s.t.} \quad &\begin{cases} \sum_{l=1}^{L_0} x_{lj} = 1 \\ a_l \leq x_{lj} \leq b_l, & l \in \omega \\ x_{lj} = 0, & l \notin \omega \end{cases} \end{aligned} \quad (45)$$

where

$$\phi_j^{(n)}(\mathbf{x}_j) \triangleq \frac{1}{2} \mathbf{x}_j' \mathbf{H} \mathbf{x}_j + \mathbf{p}' \mathbf{x}_j \quad (46)$$

$$\mathbf{H} \triangleq \mathbf{D}_j^{(n)} \quad (47)$$

$$\mathbf{p} \triangleq \nabla_{\mathbf{x}_j} \Psi(\mathbf{x}^{(n)}) - (\mathbf{x}_j^{(n)})' \mathbf{D}_j^{(n)}. \quad (48)$$

The goal of this optimization is to estimate \mathbf{x}_j and material types ω . We solve it as follows.

- 1) For each $\omega \in \Omega$, find the optimal $\hat{\mathbf{x}}_j(\omega)$ and the corresponding function values $\phi_j^{(n)}(\hat{\mathbf{x}}_j(\omega))$.

Without loss of generality, we consider the case where a given $\omega = (1, \dots, L)$ for some L between 1 and L_0 , then the optimization problem is

$$\hat{\mathbf{x}}_j(\omega) = \arg \min_{\mathbf{x}_j(\omega)} \phi_j^{(n)}(\mathbf{x}_j(\omega))$$

$$\begin{aligned} \phi_j^{(n)}(\mathbf{x}_j(\omega)) &\equiv \frac{1}{2} \mathbf{x}_j'(\omega) \mathbf{H}(\omega) \mathbf{x}_j(\omega) + \mathbf{p}'(\omega) \mathbf{x}_j(\omega) \\ \text{s.t.} \quad &\begin{cases} \sum_{l=1}^L x_{lj} = 1 \\ a_l \leq x_{lj} \leq b_l \end{cases} \end{aligned} \quad (49)$$

where $\mathbf{x}_j(\omega) \triangleq (x_{1j}, \dots, x_{Lj})$, $\mathbf{x}_j^{(n)}(\omega) \triangleq (x_{1j}^{(n)}, \dots, x_{Lj}^{(n)})$, and $\mathbf{H}(\omega)$ and $\mathbf{p}(\omega)$ are formed from elements in \mathbf{H} and \mathbf{p} with indexes corresponding to $\omega = (1, \dots, L)$ respectively.

- 2) Determine the best tuple $\hat{\omega}$ by comparing all $\phi_j^{(n)}(\hat{\mathbf{x}}_j(\omega))$, i.e.,

$$\hat{\omega} = \arg \min_{\omega \in \Omega} \phi_j^{(n)}(\hat{\mathbf{x}}_j(\omega)).$$

- 3) Obtain $\hat{\mathbf{x}}_j \equiv \hat{\mathbf{x}}_j(\hat{\omega})$ with padded zeros for $l \notin \omega$.

Given material types, i.e., given ω , the optimization problem defined in (49) is a typical convex quadratic programming problem. We used the Generalized Sequential Minimization Algorithm (GSMO) [57] to solve (49), and parallelized GSMO to update all pixels simultaneously. The pseudo-code of GSMO for solving the quadratic optimization problem with constraints in (49) is summarized in the supplementary material. One can use other quadratic programming methods to solve (49).

F. Ordered-Subset PWSQS Algorithm Outline

We use the ordered subsets approach to accelerate the ‘‘convergence’’ to a good local minimum [52] by replacing the gradient in (43) with a subset gradient scaled by the total number of subsets.

The overall ordered-subset pixel-wise separable quadratic surrogate (OS-PWSQS) algorithm for minimizing the PL cost function with constraints given in (14) is outlined in Table I.

V. RESULTS

To evaluate the proposed PL method for MMD and to compare it with the ID method [1], we simulated a DECT scan and reconstructed volume fractions of a modified NCAT chest phantom [37] containing fat, blood, omnipaque300 (iodine-based contrast agent), cortical bone, and air. We generated

TABLE I
ORDERED-SUBSET PIXEL-WISE SEPARABLE QUADRATIC SURROGATE (OS-PWSQS) ALGORITHM OUTLINE

- 1) Choose π_{mij} factors using (33).
- 2) Initialize $\mathbf{x}^{(0)}$ using the results of the image-domain method [1].
- 3) **For each iteration** $d = 1, \dots, D_{\text{iter}}$
 - a) **For each subset (subiteration)** $q = 1, \dots, Q_{\text{iter}}$
 - i) $n = d + q/Q_{\text{iter}}$
 - ii) Compute gradient of the data fidelity term \dot{L}_{qj} .

$$\dot{L}_{qj} = \sum_{i \in \mathcal{S}_q} \sum_{m=1}^{M_0} a_{mij} \nabla_{\mathbf{s}_{im}} t_{im}(\mathbf{s}_{im}^{(n)}), \quad \nabla_{\mathbf{s}_{im}} t_{im}(\mathbf{s}_{im}^{(n)}) = \nabla_{\mathbf{s}_{im}} \bar{y}_{im}(\mathbf{s}_{im}^{(n)}) - Y_{im} \nabla_{\mathbf{s}_{im}} h_{im}(\mathbf{s}_{im}^{(n)}),$$

where $\nabla_{\mathbf{s}_{im}} \bar{y}_{im}$ and $\nabla_{\mathbf{s}_{im}} h_{im}$ are given in (50) and (52) respectively, and

$$\mathbf{s}_{im}^{(n)} = \left(s_{im1}(\mathbf{x}^{(n)}), \dots, s_{imL_0}(\mathbf{x}^{(n)}) \right), \quad s_{iml}(\mathbf{x}^{(n)}) = \sum_{j=1}^{N_p} a_{mij} \mathbf{x}_i^{(n)}, \quad l = 1, \dots, L_0.$$

- iii) Compute gradient of penalty term \dot{R}_{qj} .

$$\dot{R}_{qj} = \frac{1}{Q} \left(\beta_1 \frac{\partial}{\partial x_{1j}} R_1(\mathbf{x}_1) \Big|_{\mathbf{x}_1 = \mathbf{x}_1^{(n)}}, \dots, \beta_{L_0} \frac{\partial}{\partial x_{L_0j}} R_{L_0}(\mathbf{x}_{L_0}) \Big|_{\mathbf{x}_{L_0} = \mathbf{x}_{L_0}^{(n)}} \right)$$

$$\frac{\partial}{\partial x_{lj}} R_l(\mathbf{x}_l) \Big|_{\mathbf{x}_l = \mathbf{x}_l^{(n)}} = \sum_{k \in \mathcal{N}_{lj}} \kappa_{lj} \kappa_{lk} \psi_l \left(x_{lj}^{(n)} - x_{lk}^{(n)} \right), \quad l = 1, \dots, L_0$$

- iv) Compute $L_0 \times L_0$ curvature matrices $\mathbf{D}_{qj}^{(n)}$.

$$\mathbf{D}_{qj}^{(n)} = \mathbf{D}_{\bar{L},qj}^{(n)} + \frac{1}{Q} \mathbf{D}_{R,j}^{(n)}, \quad \mathbf{D}_{\bar{L},qj}^{(n)} = \sum_{i \in \mathcal{S}_q} \sum_{m=1}^{M_0} \frac{a_{mij}^2}{\pi_{mij}} \check{\mathbf{C}}_{im}^{(n)},$$

where $\check{\mathbf{C}}_{im}^{(n)}$ and $\mathbf{D}_{R,j}^{(n)}$ are defined in (27) and (40) respectively.

- v) Compute \mathbf{H} and \mathbf{p} using (47) and (48), *i.e.*,

$$\mathbf{H} = \mathbf{D}_{qj}^{(n)}, \quad \mathbf{p} = \dot{L}_{qj} + \dot{R}_{qj} - (\mathbf{x}_j^{(n)})' \mathbf{D}_{qj}^{(n)}.$$

- vi) **For each tuple** $\omega \in \Omega$
 - A) Form $\mathbf{x}_j^{(n)}(\omega)$, $\mathbf{H}(\omega)$, $\mathbf{p}(\omega)$ by extracting elements in $\mathbf{x}_j^{(n)}$, \mathbf{H} and \mathbf{p} with indexes corresponding to ω respectively.
 - B) Obtain minimizer $\hat{\mathbf{x}}_j(\omega)$ of the QP problem in (49) using GSMO.
 - C) Compute and store minimal surrogate function value $\phi_j^{(n)}(\hat{\mathbf{x}}_j(\omega))$ using (49).

End

- vii) Determine optimal $\hat{\omega}$ by comparing all $\phi_j^{(n)}(\hat{\mathbf{x}}_j(\omega))$, *i.e.*,

$$\hat{\omega} = \arg \min_{\omega \in \Omega} \phi_j^{(n)}(\hat{\mathbf{x}}_j(\omega)).$$

- viii) Obtain $\hat{\mathbf{x}}_j \equiv \hat{\mathbf{x}}_j(\hat{\omega})$ with padded zeros for $l \notin \omega$.

- ix) Update all pixels $\mathbf{x}^{(n+1/Q_{\text{iter}})} = \hat{\mathbf{x}} = (\hat{\mathbf{x}}_1, \dots, \hat{\mathbf{x}}_j, \dots, \hat{\mathbf{x}}_{N_p})$.

End

$$\mathbf{x}^{(n+1)} = \mathbf{x}^{(n+Q_{\text{iter}}/Q_{\text{iter}})}.$$

End

virtual un-enhancement (VUE) images from the reconstructed volume fractions using these two methods.

Fig. 1 shows true volume fractions and monoenergetic image at 70 keV of the simulated NCAT chest phantom. The simulated true images were 1024×1024 and the pixel size was 0.49 mm, while the reconstructed images were 512×512 and the pixel size was 0.98 mm. We introduced this model mismatch deliberately to test the MMD methods. We down-sampled the simulated true component images to the sizes of the reconstructed images by linearly averaging, and used these down-sampled images for comparisons with the reconstructed images.

We simulated the geometry of a GE LightSpeed X-ray CT fan-beam system with an arc detector of 888 detector channels by 984 views over 360° . The size of each detector cell was 1.0239 mm. The source to detector distance was $D_{\text{sd}} = 949.075$ mm, and the source to rotation center distance was $D_{\text{s0}} = 541$ mm. We included a quarter detector offset to reduce aliasing. We used the distance-driven (DD) projector [58] to generate projections of the true object. We simulated two incident spectra of X-ray tube voltages at 140 kVp and 80 kVp, and normalized them by their corresponding total intensities by summing the intensities over all energy bins. We generated noiseless measurements \bar{y}_{im} of the simulated NCAT phantom




Fig. 1. True volume fractions and the monoenergetic image at 70 keV of the NCAT chest phantom. Volume fractions are in the range of $[0, 1]$ and the monoenergetic image is displayed over $[800, 1200]$ with the shifted HU scale where air is 0 HU and water is 1000 HU.

using (1) and the normalized spectra. To add Poisson distributed noise to the noiseless measurements \bar{y}_{im} , we first chose 2×10^5 incident photons per ray for the 140 kVp measurements, and then determined the value of incident photons per ray for the 80 kVp measurements according to the ratio of total intensities of the originally simulated spectra at high and low energies. The incident photons per ray for the 80 kVp measurements was $2 \times 10^5 \cdot I_{i2}/I_{i1} = 6 \times 10^4$ where I_{i1} and I_{i2} denote the total intensity of the i th ray for the 140 kVp and 80 kVp spectrum, respectively.

For this simulation we let the triplet material library Ω contain five triplets selected from five materials: fat, blood, omnipaque300, cortical bone, and air, excluding the combination of omnipaque300 and cortical bone and the combination of omnipaque300 and fat. (This material library is based on the fact that contrast agent does not spread into the cortical bone area and fat area.) We implemented the ID method with several different priority lists of material triplets as described in [1]. We found that the performance of the ID method depends on the priorities of material triplets in the list. We selected the priority list that produces the best ID image quality in terms of noise, artifacts and crosstalk of component images. To initialize the PL iteration, we applied a 3×3 median filter to the ID images to decrease noise, especially salt-and-pepper noise due to crosstalk among component images (see Fig. 1 in the supplementary material for the ID images.) A priority list is not used in the PL method as




Fig. 2. Fat component fraction images reconstructed by the filtered ID method (left) and the PL method (right).

it determines the optimal triplet as the one that minimizes the PWSQS function in (46) for each pixel.

We used the conventional projection-domain dual-material decomposition method with polynomial approximation [4] followed by FBP to reconstruct water-iodine density images and chose 70 and 140 keV to yield LAC pairs for the ID method. We also tried a more sophisticated dual-material decomposition method, the statistical sinogram restoration method proposed in [9], but the final reconstructed component images were very similar to those of using polynomial approximation. For the PL method we chose $\beta_l = 2^8, 2^{11}, 2^{11}, 2^8, 2^4$ and $\delta_l = 0.01, 0.01, 0.005, 0.01, 0.1$ for fat, blood, omnipaque300, cortical bone, and air, respectively. We ran 500 iterations of the optimization transfer algorithm in Table I with 41 subsets to accelerate the convergence. Because (14) is a nonconvex problem, the algorithm finds a local minimum.

Figs. 2–6 show estimated volume fractions of the five materials reconstructed by the PL method and the filtered ID method. The grayscale values represent volume fractions of each material. The big white disks in Fig. 6 are due to the circular reconstruction support. The streak-like artifacts in the reconstructed images by the filtered ID method are very similar to those in [35, Fig. 4] and [1, Fig. 6]. The PL method greatly reduces these streak-like artifacts. Material cross talk is evident in the filtered ID results. Blood went into the fat image in Fig. 2, especially in the heart region. Cortical bone presented in the blood image in the upper left image in Fig. 3. Fat appeared in the cortical bone image, as evident in the lower left image in Fig. 5. The PL method alleviated this cross-talk phenomenon very effectively. In addition, the PL method reconstructed component images with lower noise.

Fig. 4 shows the profiles of the down-sampled true and reconstructed omnipaque300 component images by the PL and filtered ID method respectively in the lower image. The locations of the profiles are indicated as a blue line and green line in the PL and filtered ID images. The PL method corrected the bias introduced by the filtered ID method. Profiles of reconstructed other component images are provided in the supplementary material.

We constructed the virtual un-enhancement (VUE) images by replacing the volume of contrast agent (omnipaque300) in each pixel by the same amount of blood according to the method introduced in [1], [35]. Fig. 7 shows true and constructed VUE images at 70 keV using the component images reconstructed by the filtered ID and PL method. The images are displayed in a window of $[800, 1200]$ with the shifted Hounsfield unit




Fig. 3. Blood component fraction images reconstructed by the filtered ID method (left) and the PL method (right). (a) FOV images. (b) ROI images.




Fig. 4. Zoomed-in omnipaque300 component fraction images reconstructed by the filtered ID method (upper left) and the PL method (upper right). Lower image shows the horizontal profiles.

(HU) scale where air is 0 HU and water is 1000 HU. Having more accurate contrast agent and blood component images, the PL method produced a VUE image that is closer to the truth, while the VUE image using the filtered ID method has more obvious residuals of contrast agent. The PL method decreased beam-hardening artifacts in the monochromatic VUE image.

We calculated the root mean square (rms) errors of the component fractions $\sqrt{(1/N_p) \sum_{j=1}^{N_p} (\hat{x}_{lj} - x_{lj})^2}$ within the reconstruct support for each material based on the down-sampled true images. Table II shows rms errors of the component images reconstructed by the ID method, filtered ID method and the PL




Fig. 5. Cortical bone component fraction images reconstructed by the filtered ID method (left) and the PL method (right). (a) Display window [0, 1]. (b) Display window [0, 0.1].




Fig. 6. Air component fraction images reconstructed by the filtered ID method (left) and the PL method (right). (a) Display window [0, 1]. (b) Display window [0, 0.1].

method. The errors of the component images were scaled by 10^3 for easy comparison. Table II also shows the rms errors of the VUE images using the ID, filtered ID and PL method. The median filtering greatly decreased the rms errors of the ID images, especially for the fat and blood basis materials. Comparing with the filtered ID method, the PL method lowered the rms errors by about 60% for fat, blood, omnipaque300, and cortical bone component images, 2% for air image, and 20% for the monochromatic VUE image at 70 keV.




Fig. 7. True and estimated VUE images at 70 KeV. Display window is [800, 1200] HU where the HU scale is shifted by 1000 HU, i.e., air is 0 HU and water is 1000 HU.

TABLE II
RMS ERROR COMPARISON OF THE RECONSTRUCTED COMPONENT FRACTION IMAGES BY THE ID METHOD, FILTERED ID METHOD AND THE PL METHOD. ERRORS OF COMPONENT FRACTION IMAGES ARE UNITLESS AND ENLARGED BY 10^3 . ERRORS OF THE VUE IMAGES ARE IN HU UNIT

Method	Fat	Blood	Omnipaque	Bone	Air	VUE
ID	156.3	134.7	4.0	38.1	49.5	54.6
Filtered ID	101.6	62.5	3.3	34.9	42.0	54.3
PL	51.3	19.4	1.4	16.6	41.2	43.6

VI. CONCLUSION

We proposed a statistical image reconstruction method for multi-material decomposition (MMD) using DECT measurements. We used a PL cost function containing a negative log-likelihood term and an edge-preserving regularization term for each basis material. We adopted the mass and volume conservation assumption and assumed each pixel contains at most three basis materials to help solve this ill-posed problem of estimating multiple sets of unknowns from two sets of sinograms. Comparing with the ID method [1], [35] that uses the same assumptions, the proposed PL method reconstructed component and monochromatic VUE images with reduced noise, streak artifacts, and cross-talk. The PL method was able to lower the rms error by about 60% for fat, blood, omnipaque300, and cortical bone basis material images, and 20% for the monochromatic VUE image at 70 keV, compared to the filtered ID method.

Due to the complexity and nonconvexity of the PL cost function it is difficult to minimize the cost function directly. We previously introduced an optimization transfer method [36] with a series of material- and pixel-wise separable quadratic surrogate functions to monotonically decrease the PL cost function. The separability in both pixel and material caused the curvature of each surrogate function to be small. The smaller the curvature, the slower the convergence rate. The constraints on each pixel couple the estimates of material fractions, so even if we used a surrogate function that is separable across materials, the minimization step would not be separable due to the constraints. In this paper, we proposed a PWSQS optimization transfer method with separable quadratic surrogate functions that decouples pixels only. The PL cost function decreases faster with the PWSQS method (results not shown). However,

iterative methods for MMD are computationally expensive. In each iteration, a MMD method requires one forward projection and one back-projection for each basis-material image, solving a constrained quadratic programming problem for each physically meaningful material tuple, and comparing results of all material tuples to determine the optimal tuple for each pixel. The simulations in this paper used 2-D fan-beam data; to apply the PWSQS method to 3-D data, more future work on accelerating the optimization process is needed. One potential accelerating method is combining OS-PWSQS with spatially nonuniform optimization transfer [59].

The PL cost function has two parameters, one regularizer coefficient β_i and one edge-preserving parameter δ_i for each material. We found that the choice of parameters for one material component influenced the reconstructed image of another component. An appropriate combination of parameters needs to be carefully determined for each application. Huh and Fessler [60] used a material-cross penalty for DECT reconstruction. This penalty used the prior knowledge that different component images have common edges; this idea could be used for MMD as well. Choosing regularizers for the PL method and optimizing the parameters needs further investigation.

Since the PL cost function is nonconvex, good initialization is important. We used the results of the ID method followed by median filtering as the initialization of the PL method. The median filtering decreased noise and rms errors of the ID images, but did not preserve the sum-to-one constraint that the ID images satisfied. Even with this initialization the PL method was able to converge to a good local minimum that is close to the truth. As future work we will investigate image-domain “statistical” reconstruction methods that are computationally more practical than the PL method. Such methods could also serve to initialize the PL method.

We used the contrast agent, omnipaque300, as a basis material in this paper. Alternatively one can use diluted contrast agent or iodine as basis materials. Future work would investigate the effects of using various basis materials. Future work also includes applying the PL method to real spectral CT data, e.g., from fast-kVp switching DE scans, dual-source scans, or dual-layer detectors, to decompose materials as needed by the application.

APPENDIX
SURROGATE FUNCTION DERIVATIONS

This section describes derivation details of the surrogate design in Section IV.

A. Convexity Proof

This section proves the convexity of $h_{im}(\mathbf{s}_{im})$. The gradient of $\bar{y}_{im}(\mathbf{s}_{im})$ with respect to \mathbf{s}_{im} is

$$\nabla_{\mathbf{s}_{im}} \bar{y}_{im}(\mathbf{s}_{im}) = - \int I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}(\mathbf{x})} \boldsymbol{\mu}(\mathcal{E}) d\mathcal{E} \quad (50)$$

and the Hessian is

$$\nabla_{\mathbf{s}_{im}}^2 \bar{y}_{im}(\mathbf{s}_{im}) = \int I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}(\mathbf{x})} \boldsymbol{\mu}(\mathcal{E}) \boldsymbol{\mu}'(\mathcal{E}) d\mathcal{E}. \quad (51)$$

Since the Hessian matrix is positive-semidefinite, *i.e.*, $\nabla_{\mathbf{s}_{im}}^2 \bar{y}_{im}(\mathbf{s}_{im}) \succeq \mathbf{0}$, the function $\bar{y}_{im}(\mathbf{s}_{im})$ is convex.

The gradient of $h_{im}(\mathbf{s}_{im})$ with respect to \mathbf{s}_{im} is

$$\nabla_{\mathbf{s}_{im}} h_{im}(\mathbf{s}_{im}) = \frac{1}{\bar{y}_{im}(\mathbf{s}_{im})} \nabla_{\mathbf{s}_{im}} \bar{y}_{im}(\mathbf{s}_{im}). \quad (52)$$

The Hessian of $h_{im}(\mathbf{s}_{im})$ is

$$\begin{aligned} \nabla_{\mathbf{s}_{im}}^2 h_{im}(\mathbf{s}_{im}) &= \frac{1}{\bar{y}_{im}(\mathbf{s}_{im})} \int I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}(\mathbf{x})} \boldsymbol{\mu}(\mathcal{E}) \boldsymbol{\mu}'(\mathcal{E}) d\mathcal{E} \\ &\quad - \frac{1}{\bar{y}_{im}^2(\mathbf{s}_{im})} \left(\int I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}(\mathbf{x})} \boldsymbol{\mu}(\mathcal{E}) d\mathcal{E} \right) \\ &\quad \cdot \left(\int I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}(\mathbf{x})} \boldsymbol{\mu}(\mathcal{E}) d\mathcal{E} \right)' \\ &= \frac{z_{im}(\mathbf{s}_{im})}{\bar{y}_{im}(\mathbf{s}_{im})} \int q_{im}(\mathcal{E}) \boldsymbol{\mu}(\mathcal{E}) \boldsymbol{\mu}'(\mathcal{E}) d\mathcal{E} \\ &\quad - \frac{z_{im}^2(\mathbf{s}_{im})}{\bar{y}_{im}^2(\mathbf{s}_{im})} \left(\int q_{im}(\mathcal{E}) \boldsymbol{\mu}(\mathcal{E}) d\mathcal{E} \right) \\ &\quad \cdot \left(\int q_{im}(\mathcal{E}) \boldsymbol{\mu}(\mathcal{E}) d\mathcal{E} \right)' \\ &\succeq \frac{z_{im}^2(\mathbf{s}_{im})}{\bar{y}_{im}^2(\mathbf{s}_{im})} \int q_{im}(\mathcal{E}) \boldsymbol{\mu}(\mathcal{E}) \boldsymbol{\mu}'(\mathcal{E}) d\mathcal{E} \\ &\quad - \frac{z_{im}^2(\mathbf{s}_{im})}{\bar{y}_{im}^2(\mathbf{s}_{im})} \left(\int q_{im}(\mathcal{E}) \boldsymbol{\mu}(\mathcal{E}) d\mathcal{E} \right) \\ &\quad \cdot \left(\int q_{im}(\mathcal{E}) \boldsymbol{\mu}(\mathcal{E}) d\mathcal{E} \right)' \\ &= \frac{z_{im}^2(\mathbf{s}_{im})}{\bar{y}_{im}^2(\mathbf{s}_{im})} \text{Cov} \{ \boldsymbol{\mu}(\mathcal{E}) \} \\ &\succeq \mathbf{0} \end{aligned} \quad (53)$$

where we define the following ‘‘probability density function’’

$$q_{im}(\mathcal{E}) \triangleq \frac{I_{im}(\mathcal{E}) e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}(\mathbf{x})}}{z_{im}(\mathbf{s}_{im})} \quad (54)$$

and used the fact that $z_{im}(\mathbf{s}_{im}) \leq \bar{y}_{im}(\mathbf{s}_{im})$ for the first inequality. Since the Hessian matrix of $h_{im}(\mathbf{s}_{im})$ is positive-semidefinite, $h_{im}(\mathbf{s}_{im})$ is a convex function of \mathbf{s}_{im} .

B. Derivation of Nonseparable Quadratic Surrogate

The section derives the nonseparable quadratic surrogate in Section IV-B2 in details.

The optimal curvature $\check{c}(\cdot; \alpha)$ for $f(x; \alpha)$ derived in [52] is

$$\check{c}(x; \alpha) = \begin{cases} \left[2 \frac{1 - e^{-x} - x e^{-x}}{x^2} \right]_+, & x > 0 \\ 1 & x = 0. \end{cases} \quad (55)$$

Thus, the optimal quadratic surrogate for $f(x; \alpha)$ is

$$q(x; x_0, \alpha) \triangleq f(x_0; \alpha) + \dot{f}(x_0; \alpha)(x - x_0) + \frac{1}{2} \check{c}(x_0; \alpha)(x - x_0)^2, \quad (56)$$

where $q(x; x_0, \alpha) \geq f(x; \alpha)$.

Substituting (56) into (24) leads to a nonseparable quadratic surrogate $f_{2,im}^{(n)}(\mathbf{s}_{im})$ of $f_{1,im}^{(n)}(\mathbf{s}_{im})$, *i.e.*,

$$\begin{aligned} f_{2,im}^{(n)}(\mathbf{s}_{im}) &\triangleq C + \int I_{im}(\mathcal{E}) \left(f \left(\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}; Y_{im}^{(n)} e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}} \right) \right. \\ &\quad \left. + \dot{f} \left(\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}; Y_{im}^{(n)} e^{-\boldsymbol{\mu}(\mathcal{E}) \cdot \mathbf{s}_{im}^{(n)}} \right) \boldsymbol{\mu}(\mathcal{E}) \cdot \left(\mathbf{s}_{im} - \mathbf{s}_{im}^{(n)} \right) \right. \\ &\quad \left. + \frac{1}{2} \check{c}_{im}^{(n)}(\mathcal{E}) \left(\boldsymbol{\mu}(\mathcal{E}) \cdot \left(\mathbf{s}_{im} - \mathbf{s}_{im}^{(n)} \right) \right)^2 \right) d\mathcal{E} \end{aligned} \quad (57)$$

where $\check{c}_{im}^{(n)}(\mathcal{E})$ is defined in (28). Because $f_{2,im}^{(n)}(\mathbf{s}_{im})$ is a quadratic function of \mathbf{s}_{im} , one can rewrite it as (26).

C. Derivation of Pixel-Wise Separable Quadratic Surrogate

This section derives the pixel-wise separable quadratic surrogate in Section IV-B3.

We rewrite the sinogram vector as

$$\begin{aligned} \mathbf{s}_{im}(\mathbf{x}) &= \sum_{j=1}^{N_p} a_{mij} \mathbf{x}_j \\ &= \sum_{j=1}^{N_p} \pi_{mij} \left(\frac{a_{mij}}{\pi_{mij}} (\mathbf{x}_j - \mathbf{x}_j^{(n)}) + \sum_{j=1}^{N_p} a_{mij} \mathbf{x}_j^{(n)} \right) \end{aligned} \quad (58)$$

provided $\sum_{j=1}^{N_p} \pi_{mij} = 1$ and π_{mij} is zero only if a_{mij} is zero. If the π_{mij} 's are nonnegative, then we can apply the convexity inequality to the quadratic function $f_{2,im}^{(n)}(\mathbf{s}_{im}(\mathbf{x}))$ defined in (26) to write

$$\begin{aligned} f_{2,im}^{(n)}(\mathbf{s}_{im}(\mathbf{x})) &= f_{2,im}^{(n)} \left(\sum_{j=1}^{N_p} \pi_{mij} \left(\frac{a_{mij}}{\pi_{mij}} (\mathbf{x}_j - \mathbf{x}_j^{(n)}) + \sum_{j=1}^{N_p} a_{mij} \mathbf{x}_j^{(n)} \right) \right) \\ &\leq \sum_{j=1}^{N_p} \pi_{mij} f_{2,im}^{(n)}(b_{mij}(\mathbf{x}_j)) \end{aligned} \quad (59)$$

where b_{mij} is defined in (32).

The value of $L_3^{(n)}$ evaluated at $\mathbf{x}^{(n)}$ is

$$\begin{aligned} L_3^{(n)}(\mathbf{x}) \Big|_{\mathbf{x}=\mathbf{x}^{(n)}} &= \sum_{j=1}^{N_p} f_{3,j}^{(n)}(\mathbf{x}_j^{(n)}) \\ &= \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} f_{2,im}^{(n)}(\mathbf{s}_{im}(\mathbf{x}^{(n)})) \\ &= \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} t_{im}(\mathbf{s}_{im}(\mathbf{x}^{(n)})) \\ &= \bar{L}(\mathbf{x}) \Big|_{\mathbf{x}=\mathbf{x}^{(n)}}. \end{aligned} \quad (60)$$

The column gradient of $L_3^{(n)}$ has elements x_{lj}

$$\begin{aligned} \frac{\partial}{\partial x_{lj}} L_3^{(n)}(\mathbf{x}) \Big|_{\mathbf{x}_j=\mathbf{x}_j^{(n)}} &= \frac{\partial}{\partial x_{lj}} f_{3,j}^{(n)}(\mathbf{x}_j) \Big|_{\mathbf{x}_j=\mathbf{x}_j^{(n)}} \\ &= \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} a_{mij} \dot{f}_{2,im}^{(n)} \left(\sum_{j=1}^{N_p} a_{mij} \mathbf{x}_j^{(n)} \right) \\ &= \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} a_{mij} \dot{t}_{im} \left(\sum_{j=1}^{N_p} a_{mij} \mathbf{x}_j^{(n)} \right) \\ &= \frac{\partial}{\partial x_{lj}} \bar{L}(\mathbf{x}) \Big|_{\mathbf{x}_j=\mathbf{x}_j^{(n)}}. \end{aligned} \quad (61)$$

The ‘‘matched function value’’ and ‘‘matched derivative’’ properties are inherent to optimization transfer methods [51].

The Hessian $\mathbf{D}_{\bar{L},j}^{(n)} \triangleq \nabla^2 f_{3,j}^{(n)}(\mathbf{x})$ has elements

$$\begin{aligned} [\mathbf{D}_{\bar{L},j}^{(n)}]_{lk} &= \frac{\partial^2}{\partial x_{lj} \partial x_{kj}} f_{3,j}^{(n)}(\mathbf{x}_j) \\ &= \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} a_{mij} \frac{\partial^2}{\partial x_{lj} \partial x_{kj}} f_{2,im}^{(n)}(b_{mij}(\mathbf{x}_j)) \\ &= \sum_{i=1}^{N_d} \sum_{m=1}^{M_0} \frac{a_{mij}^2}{\pi_{mij}} [\mathbf{C}_{im}^{(n)}]_{lk} \end{aligned} \quad (62)$$

where $\mathbf{C}_{im}^{(n)}$ is defined in (28).

ACKNOWLEDGMENT

The authors would like to thank P. Mendonca for discussions about the ID method. J. Fessler would like to thank P. Kinahan and A. Alessio for numerous discussions related to DE imaging.

REFERENCES

- [1] P. R. S. Mendonca, P. Lamb, and D. Sahani, ‘‘A flexible method for multi-material decomposition of dual-energy CT images,’’ *IEEE Trans. Med. Imag.*, vol. 33, no. 1, pp. 99–116, Jan. 2014.
- [2] I. A. Elbakri and J. A. Fessler, ‘‘Segmentation-free statistical image reconstruction for polyenergetic X-ray computed tomography with experimental validation,’’ *Phys. Med. Biol.*, vol. 48, no. 15, pp. 2543–2578, Aug. 2003.

- [3] R. A. Brooks and G. Di Chiro, ‘‘Beam-hardening in X-ray reconstruction tomography,’’ *Phys. Med. Biol.*, vol. 21, no. 3, pp. 390–398, May 1976.
- [4] R. E. Alvarez and A. Macovski, ‘‘Energy-selective reconstructions in X-ray computed tomography,’’ *Phys. Med. Biol.*, vol. 21, no. 5, pp. 733–744, Sep. 1976.
- [5] P. Sukovic and N. H. Clinthorne, ‘‘Basis material decomposition using triple-energy X-ray computed tomography,’’ in *Proc. IEEE Instrum. Meas. Technol. Conf.*, Venice, Italy, 1999, vol. 3, pp. 1615–1618.
- [6] P. Sukovic and N. H. Clinthorne, ‘‘Penalized weighted least-squares image reconstruction for dual energy X-ray transmission tomography,’’ *IEEE Trans. Med. Imag.*, vol. 19, no. 11, pp. 1075–1081, Nov. 2000.
- [7] P. E. Kinahan, A. M. Alessio, and J. A. Fessler, ‘‘Dual energy CT attenuation correction methods for quantitative assessment of response to cancer therapy with PET/CT imaging,’’ *Technol. Cancer Res. Treatment*, vol. 5, no. 4, pp. 319–328, Aug. 2006.
- [8] L.-J. Zhang, J. Peng, S.-Y. Wu, Z. Jane Wang, X.-S. Wu, C.-S. Zhou, X.-M. Ji, and G.-M. Lu, ‘‘Liver virtual non-enhanced CT with dual-source, dual-energy CT: A preliminary study,’’ *Eur. Radiol.*, vol. 20, no. 9, pp. 2257–2264, 2010.
- [9] J. Noh, J. A. Fessler, and P. E. Kinahan, ‘‘Statistical sinogram restoration in dual-energy CT for PET attenuation correction,’’ *IEEE Trans. Med. Imag.*, vol. 28, no. 11, pp. 1688–1702, Nov. 2009.
- [10] A. J. Coleman and M. Sinclair, ‘‘A beam-hardening correction using dual-energy computed tomography,’’ *Phys. Med. Biol.*, vol. 30, no. 11, pp. 1251–1256, Nov. 1985.
- [11] M. M. Goodsitt, ‘‘Beam hardening errors in post-processing dual energy quantitative computed tomography,’’ *Med. Phys.*, vol. 22, no. 7, pp. 1039–1047, Jul. 1995.
- [12] A. Macovski, R. E. Alvarez, J. Chan, J. P. Stonestrom, and L. M. Zatz, ‘‘Energy dependent reconstruction in X-ray computerized tomography,’’ *Comput. Biol. Med.*, vol. 6, no. 4, pp. 325–336, Oct. 1976.
- [13] R. E. Alvarez and A. Macovski, ‘‘X-ray spectral decomposition imaging system,’’ U.S. Patent 4 029 963, Jun. 14, 1977.
- [14] W. H. Marshall, R. E. Alvarez, and A. Macovski, ‘‘Initial results with pre-reconstruction dual-energy computed tomography (PREDECT),’’ *Radiol.*, vol. 140, no. 2, pp. 421–430, Aug. 1981.
- [15] J. P. Stonestrom, R. E. Alvarez, and A. Macovski, ‘‘A framework for spectral artifact corrections in X-ray CT,’’ *IEEE Trans. Biomed. Eng.*, vol. 28, no. 2, pp. 128–141, Feb. 1981.
- [16] W. Huh, J. A. Fessler, A. M. Alessio, and P. E. Kinahan, ‘‘Fast kVp-switching dual energy CT for PET attenuation correction,’’ in *Proc. IEEE Nucl. Sci. Symp. Med. Imag. Conf.*, 2009, pp. 2510–2515.
- [17] W. Huh and J. A. Fessler, ‘‘Model-based image reconstruction for dual-energy X-ray CT with fast kVp switching,’’ in *Proc. IEEE Int. Symp. Biomed. Imag.*, 2009, pp. 326–329.
- [18] A. N. Primak, J. C. Ramirez Giraldo, X. Liu, L. Yu, and C. H. McCollough, ‘‘Improved dual-energy material discrimination for dual-source CT by means of additional spectral filtration,’’ *Med. Phys.*, vol. 36, no. 4, pp. 1359–1369, Mar. 2009.
- [19] G. C. Giakos, S. Chowdhury, N. Shah, K. Mehta, S. Sumrain, A. Passerini, N. Patnekar, E. A. Evans, L. Fraiwan, O. C. Ugweje, and R. Nemer, ‘‘Signal evaluation of a novel dual-energy multimedia imaging sensor,’’ *IEEE Trans. Med. Imag.*, vol. 51, no. 5, pp. 949–954, Oct. 2002.
- [20] S. Kappler, M. Grasruck, D. Niederlöhner, M. Strassburg, and S. Wirth, ‘‘Dual-energy performance of dual-kVp in comparison to dual-layer and quantum-counting CT system concepts,’’ in *Proc. SPIE Med. Imag. Phys. Med. Imag.*, 2009, p. 725842.
- [21] L. Yu, X. Liu, and C. H. McCollough, ‘‘Pre-reconstruction three-material decomposition in dual-energy CT,’’ in *Proc. SPIE Med. Imag. Phys. Med. Imag.*, 2009, p. 72581V.
- [22] X. Liu, L. Yu, A. N. Primak, and C. H. McCollough, ‘‘Quantitative imaging of element composition and mass fraction using dual-energy CT: Three-material decomposition,’’ *Med. Phys.*, vol. 36, no. 5, pp. 1602–1609, Apr. 2009.
- [23] L. P. Nogueira, R. C. Barroso, C. J. G. Pinheiro, D. Braz, L. F. de Oliveira, G. Tromba, and N. Sodini, ‘‘Mapping lead distribution in bones by dual-energy computed microtomography with synchrotron radiation,’’ in *Proc. IEEE Nucl. Sci. Symp. Med. Imag. Conf.*, 2009, pp. 3471–3474.
- [24] C. Maass, R. Grimmer, and M. Kachelriess, ‘‘Dual energy CT material decomposition from inconsistent rays (MDIR),’’ in *Proc. IEEE Nucl. Sci. Symp. Med. Imag. Conf.*, 2009, pp. 3446–3452.

- [25] C. Cai, T. Rodet, S. Legoupil, and A. Mohammad-Djafari, "A full-spectral Bayesian reconstruction approach based on the material decomposition model applied in dual-energy computed tomography," *Med. Phys.*, vol. 40, no. 11, p. 111916, Nov. 2013.
- [26] Y. Jin, G. Fu, V. Lobastov, P. M. Edic, and B. De Man, "Dual-energy performance of x-ray CT with an energy-resolved photon-counting detector in comparison to x-ray CT with dual kVp," in *Proc. Int. Meet. Fully 3-D Image Reconstruct. Radiol. Nucl. Med.*, 2013, pp. 456–460.
- [27] R. T. Ritchings and B. R. Pullan, "A technique for simultaneous dual energy scanning," *J. Comp. Assist. Tomogr.*, vol. 3, no. 6, pp. 842–846, Dec. 1979.
- [28] R. Taschereau, R. W. Silverman, and A. F. Chatziioannou, "Dual-energy attenuation coefficient decomposition with differential filtration and application to a microCT scanner," *Phys. Med. Biol.*, vol. 55, no. 4, pp. 1141–1155, Feb. 2010.
- [29] Y. Long, J. A. Fessler, and J. M. Balter, "Two-material decomposition from a single CT scan using statistical image reconstruction," in *Proc. Am. Assoc. Phys. Med.*, 2011.
- [30] B. Rutt and A. Fenster, "Split-filter computed tomography: A simple technique for dual energy scanning," *J. Comp. Assist. Tomogr.*, vol. 4, no. 4, pp. 501–509, Aug. 1980.
- [31] A. D. Laidevant, S. Malkov, C. I. Flowers, K. Kerlikowske, and J. A. Shepherd, "Compositional breast imaging using a dual-energy mammography protocol," *Med. Phys.*, vol. 37, no. 1, pp. 164–174, Jan. 2010.
- [32] B. J. Kis, Z. Sarnyai, R. Kakonyi, M. Erdelyi, and G. Szabo, "Fat and iron quantification of the liver with dual-energy computed tomography in the presence of high atomic number elements," in *Proc. 2nd Int. Meet. Image Format. X-Ray CT*, 2012, pp. 230–233.
- [33] S. Masetti, L. Roma, P. L. Rossi, N. Lanconelli, and G. Baldazzi, "Preliminary results of a multi-energy CT system for small animals," *J. Instrum.*, vol. 4, no. 6, p. 06011, Jun. 2009.
- [34] X. Wang, D. Meier, S. Mikkelsen, G. E. Maehlum, D. J. Wagenaar, B. M. W. Tsui, B. E. Patt, and E. C. Frey, "MicroCT with energy-resolved photon-counting detectors," *Phys. Med. Biol.*, vol. 56, no. 9, pp. 2791–2816, May 2011.
- [35] P. R. S. Mendonca, R. Bhotika, B. W. Thomsen, P. E. Licato, and M. C. Joshi, "Multi-material decomposition of dual-energy CT data," in *Proc. SPIE*, 2010, vol. 7622, p. 76221W.
- [36] Y. Long and J. A. Fessler, "Multi-material decomposition using statistical image reconstruction in X-ray CT," in *Proc. 2nd Int. Meet. Image Format. X-Ray CT*, 2012, pp. 413–416.
- [37] W. P. Segars and B. M. W. Tsui, "Study of the efficacy of respiratory gating in myocardial SPECT using the new 4-D NCAT phantom," *IEEE Trans. Nucl. Sci.*, vol. 49, no. 3, pp. 675–679, Jun. 2002.
- [38] C. Ruth and P. M. Joseph, "Estimation of a photon energy spectrum for a computed tomography scanner," *Med. Phys.*, vol. 24, no. 5, pp. 695–702, May 1997.
- [39] A. P. Colijn and F. J. Beekman, "Accelerated simulation of cone beam X-ray scatter projections," *IEEE Trans. Med. Imag.*, vol. 23, no. 5, pp. 584–590, May 2004.
- [40] M. Yavuz and J. A. Fessler, "Statistical image reconstruction methods for randoms-precorrected PET scans," *Med. Image Anal.*, vol. 2, no. 4, pp. 369–378, Dec. 1998.
- [41] M. Yavuz and J. A. Fessler, "Penalized-likelihood estimators and noise analysis for randoms-precorrected PET transmission scans," *IEEE Trans. Med. Imag.*, vol. 18, no. 8, pp. 665–674, Aug. 1999.
- [42] P. Sukovic and N. H. Clinthorne, "Data weighted vs. non-data weighted dual energy reconstructions for X-ray tomography," in *Proc. IEEE Nucl. Sci. Symp. Med. Imag. Conf.*, 1998, vol. 3, pp. 1481–1483.
- [43] R. Zhang, J. B. Thibault, C. A. Bouman, K. D. Sauer, and J. Hsieh, "Model-based iterative reconstruction for dual-energy X-ray CT using a joint quadratic likelihood model," *IEEE Trans. Med. Imag.*, vol. 33, no. 1, pp. 117–134, Jan. 2014.
- [44] G. H. Glover and N. J. Pelc, "Nonlinear partial volume artifacts in x-ray computed tomography," *Med. Phys.*, vol. 7, no. 3, pp. 238–248, May 1980.
- [45] P. M. Joseph and R. D. Spital, "The exponential edge-gradient effect in x-ray computed tomography," *Phys. Med. Biol.*, vol. 26, no. 3, pp. 473–487, May 1981.
- [46] J. A. Fessler and W. L. Rogers, "Spatial resolution properties of penalized-likelihood image reconstruction methods: Space-invariant tomographs," *IEEE Trans. Image Process.*, vol. 5, no. 9, pp. 1346–1358, Sep. 1996.
- [47] J. Valenzuela and J. A. Fessler, "Joint reconstruction of Stokes images from polarimetric measurements," *J. Opt. Soc. Am. A*, vol. 26, no. 4, pp. 962–968, Apr. 2009.
- [48] X. He, J. A. Fessler, L. Cheng, and E. C. Frey, "Regularized image reconstruction algorithms for dual-isotope myocardial perfusion SPECT (MPS) imaging using a cross-tracer edge-preserving prior," *IEEE Trans. Med. Imag.*, vol. 30, no. 6, pp. 1169–1183, Jun. 2011.
- [49] K. Lange, D. R. Hunter, and I. Yang, "Optimization transfer using surrogate objective functions," *J. Computat. Graph. Stat.*, vol. 9, no. 1, pp. 1–20, Mar. 2000.
- [50] A. R. De Pierro, "A modified expectation maximization algorithm for penalized likelihood estimation in emission tomography," *IEEE Trans. Med. Imag.*, vol. 14, no. 1, pp. 132–137, Mar. 1995.
- [51] M. W. Jacobson and J. A. Fessler, "An expanded theoretical treatment of iteration-dependent majorize-minimize algorithms," *IEEE Trans. Image Process.*, vol. 16, no. 10, pp. 2411–2422, Oct. 2007.
- [52] H. Erdoğan and J. A. Fessler, "Monotonic algorithms for transmission tomography," *IEEE Trans. Med. Imag.*, vol. 18, no. 9, pp. 801–814, Sep. 1999.
- [53] H. Erdoğan and J. A. Fessler, "Ordered subsets algorithms for transmission tomography," *Phys. Med. Biol.*, vol. 44, no. 11, pp. 2835–2851, Nov. 1999.
- [54] H. M. Hudson and R. S. Larkin, "Accelerated image reconstruction using ordered subsets of projection data," *IEEE Trans. Med. Imag.*, vol. 13, no. 4, pp. 601–609, Dec. 1994.
- [55] J. A. Fessler, "Statistical image reconstruction methods for transmission tomography," in *Handbook of Medical Imaging, Volume 2. Medical Image Processing and Analysis*, M. Sonka and J. M. Fitzpatrick, Eds. Bellingham, WA: SPIE, 2000, pp. 1–70.
- [56] P. J. Huber, *Robust Statistics*. New York: Wiley, 1981.
- [57] S. S. Keerthi and E. G. Gilbert, "Convergence of a generalized SMO algorithm for SVM classifier design," *Mach. Learn.*, vol. 46, no. 1–3, pp. 351–360, 2002.
- [58] B. De Man and S. Basu, "Distance-driven projection and backprojection in three dimensions," *Phys. Med. Biol.*, vol. 49, no. 11, pp. 2463–2475, Jun. 2004.
- [59] D. Kim, D. Pal, J.-B. Thibault, and J. A. Fessler, "Accelerating ordered subsets image reconstruction for X-ray CT using spatially non-uniform optimization transfer," *IEEE Trans. Med. Imag.*, vol. 32, no. 11, pp. 1965–1978, Nov. 2013.
- [60] W. Huh and J. A. Fessler, "Iterative image reconstruction for dual-energy x-ray CT using regularized material sinogram estimates," in *Proc. IEEE Int. Symp. Biomed. Imag.*, 2011, pp. 1512–1515.