


Reducing short-scan artifacts in 3D axial cone-beam CT with extra views (Mo21)

J H Cho, D Pal, J B Thibault, J A Fessler
University of Michigan and GE Healthcare Technologies



- ▶ Short-scans have better temporal resolution in CT
- ▶ Extra views (beyond short scan) can improve IQ in poorly sampled regions
- ▶ Main Idea: use information in extra views via PWLS to improve IQ in poorly sampled regions while preserving temporal resolution elsewhere.

